



CONVERSATION RECORD

05/16/2014

NAME OF PERSON(S) CONTACTED OR IN CONTACT WITH YOU Pete Auksel		DATE OF CONTACT 05/16/2014	TYPE OF CONVERSATION <input type="checkbox"/> E-MAIL <input checked="" type="checkbox"/> TELEPHONE <input type="checkbox"/> INCOMING <input checked="" type="checkbox"/> OUTGOING
E-MAIL ADDRESS pjauksel@uss.com		TELEPHONE NUMBER (219) 888-7841	

ORGANIZATION Radiation Safety Officer US Steel Corporation	DOCKET NUMBER(S) 030-29318; 030-36022; 030-18249; 030-29080
--	--

LICENSE NUMBER(S) 13-23600-01; 13-26104-03; 13-07964-07; 13-07964-08	CONTROL NUMBER(S) 583352; 583353; 583484; 583467
---	---

SUBJECT
Additional Information Needed

SUMMARY
In you most requests, you have asked for several licenses to be combined. Additional information will be needed to complete this request.

1. For NRC License 13-23600-01, you have requested to add the materials from your General License No. 705350-15. However, you have not provided any leak tests for the devices you wish to transfer to this license. Please provide the most current leak tests for all gauges listed under General License No. 705350-15. For any gauges that have been returned to the manufacturer, please provide the leak tests.
2. For NRC License 13-07964-07, you have requested to combine this license with NRC License 13-26104-03 and for this license to be terminated. However, you have not provided any leak tests for the devices you wish to transfer to this license. Please provide the most current leak tests for all gauges listed under NRC License 13-07964-07.


Continue on Page 2

ACTION REQUIRED (IF ANY)
Please submit your response by May 23, 2014, and reference it to my attention as "additional information to control numbers 583352; 583353; 583484; 583467" to facilitate proper handling in our office. Your response must be currently dated and signed by a senior management representative. If you have any questions or require clarification of any of the information stated above, please do not hesitate to contact me at 630-829-9607

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter and its enclosure will be available electronically for public inspection in the NRC Public Document Room or from the NRC's Agencywide Documents Access and Management System (ADAMS), accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>.

Continue on Page 3

NAME OF PERSON DOCUMENTING CONVERSATION
Jennifer L. Bishop

SIGNATURE
 5/16/14

CONVERSATION RECORD (continued)

SUMMARY: (Continued from page 1)

3. For NRC License 13-07964-08, you have requested to combine this license with NRC License 13-26104-03 and for this license to be terminated. However, two of the leak tests that were provided are out of date. Please provide the most current leak tests for the gauges with serial numbers 19701 and 19703.

4. While reviewing the documentation you provided, we noticed that you provided inventory sheets and documentation for gauges returned to the manufacturer that are for licenses where no changes were requested. Please note that at this time we will not be taking any action on any licenses for which you have not specifically requested changes.