

From: Tripp, Lester
Sent: Thursday, July 24, 2014 9:46 AM
To: rayacarlson@att.net
Subject: Exposure Limits in unrestricted areas Eranga Cardiology

Licensee: Eranga Cardiology
License No. 07-35021-01
Docket No. 03038592
Control No. 584042

Dear Mr. Carlson:

Please confirm that your facility will be in compliance with 10 CFR 20.1301. Specifically, the no member of the public will receive in excess of 2.0 mrem in any one hour, and 100 mrem in any year.

Thank you

Lester