

From: [Lanzisera, Penny](mailto:Lanzisera.Penny)
To: ralph.sgambato@greenwichhospital.org
Subject: Request for additional information to add an Authorized User and Authorized Medical Physicist
Date: Friday, August 01, 2014 2:41:00 PM

Greenwich Hospital Association
L.N. 06-09522-01
D.N. 03001276
Mail Control 584057

Mr. Sgambato,

In order to complete our review to add Dr. Johung, we have requested the following information from Yale New Haven Hospital (YNHH):

1. Does Dr. Johung have both an MD and PhD?
2. On March 27, 2014, you prepared a memo documenting that Dr. Johung was approved by the YNHH RSC for 35.600 remote afterloading use; however a review of NRC Form 313A signed by Dr. Roy Decker on September 17, 2013, only documents 300 hours of supervised work experience that occurred during the periods of July 2009 to June 2013. Was this an error, since 500 hours of supervised work experience is required? If so, please submit a statement from Dr. Decker indicating that 500 hours was completed by Dr. Johung.
3. Please confirm that Dr. Decker is an AU for remote afterloaders at YNHH.
4. Was Dr. Johung approved at YNHH for Radium-223 use and if so, please provide the documentation.

We have no further questions with respect to the addition of Niranjan Bhandare, M.S. as an authorized medical physicist.

Discussions with Yale indicate that the information will be forthcoming; however, if we do not receive a reply within 30 days we will process an amendment to add the physicist to the license only. Please contact me directly if you have any questions.

Penny Lanzisera
Senior Health Physicist
US NRC, Region I