

## Hill, Carol

---

**From:** Simmons, Michelle  
**Sent:** Tuesday, September 16, 2014 4:35 PM  
**To:** Hill, Carol  
**Subject:** FW: Bonner General Health: License Renewal Deficiency Letter

Carol,

Please add this to ADAMS.  
Non-sensitive, Public

---

**From:** Simmons, Michelle  
**Sent:** Tuesday, September 16, 2014 4:34 PM  
**To:** 'maileonmeadow@yahoo.com'  
**Subject:** Bonner General Health: License Renewal Deficiency Letter

Docket No.: 03036658  
License No. : 11-27785-01  
Mail Control No.: 584302

This is in reference to the application dated July 3, 2014, requesting the renewal of NRC license number 11-27785-01. In order to continue our review, we need the following additional information.

Please make the following commitments:

1. Radiation monitoring instruments will be calibrated by a person qualified to perform survey meter calibrations.
2. We reserve the right to upgrade our survey instruments as necessary as long as they are adequate to measure the type and level of radiation for which they are used.
3. Equipment used to measure dosages will be calibrated in accordance with nationally recognized standards or the manufacturer's instructions.
4. Either we will perform a prospective evaluation demonstrating that unmonitored individuals are not likely to receive, in one year, a radiation dose in excess of 10% of the allowable limits in 10 CFR Part 20 or we will provide dosimetry that meets the requirements listed under Criteria in NUREG- 1556, Vol. 9, Consolidated Guidance About Materials Licenses: Program-Specific Guidance About Medical Use Licensees, dated October 2002.
5. We have developed and will implement and maintain written waste disposal procedures for licensed material in accordance with 10 CFR 20.1101, that also meet the requirements of the applicable section of Subpart K to 10 CFR Part 20 and 10 CFR 35.92
6. A statement that: We have developed and will implement and maintain written procedures for safe response to spills of licensed material in accordance with 10 CFR 20.1101.

Please provide this information by **September 23, 2014**. You may fax your signed response to 817-200-1263, referencing mail control 584302. When you send the fax, you may wish to leave a voicemail or e-mail message to alert me to look for it. If you are submitting your response by email, the response must be submitted in pdf format. Please send me an e-mail or call me at 817-200-1590 if you have any questions.

Michelle Simmons  
Health Physicist  
US Nuclear Regulatory Commission  
1600 East Lamar Blvd.  
Arlington, Texas 76011  
817-200-1590