



Docket No: 03036658
License No: 11-27785-01
Mail control No: 584302
9/22/2014

This is in reference to our application dated July 3, 2014 submitting additional information requested.

- Item 1. We commit to having our monitoring instruments calibrated annually by a person *qualified* to perform survey meter calibrations.
- Item 2. We reserve the right to upgrade our survey instruments as necessary as long as they are adequate to measure the type and level of radiation for which they are used.
- Item 3. We commit to have equipment used to measure dosages calibrated in accordance with nationally recognized standards or the manufacturer's instructions.
- Item 4. We will provide dosimetry that meets the requirements listed under criteria in NUREG 1556, vol 9 October 2002 Program Specific Guidance About Medical Licensees.
- Item 5. We will implement and maintain written waste disposal procedures for licensed material in accordance with 10CFR20.1101, that also meet the requirements of the applicable section of Subpart K to 10 CFR part 20 and 10 CFR 35.92.
- Item 6. We have developed and will implement and maintain written procedures for safe response to spills of licensed material in accordance with 10 CFR 20.1101.

PUBLIC

- Immediate Release
 Normal Release

NON-PUBLIC

- A.3 Sensitive-Security Related
 A.7 Sensitive Internal
 Other:

Reviewer: MS

Date: 9/22/14

Jami Hicks, CMR

FAX

To: Michelle Simmer From: Tami Hides

Date: 9/22/14 Pgs. (including cover) 2

Re: Mail control 584302, lic # 11-27785-01

Michelle -

I emailed doc
as well.

Thank you for
your time.

CONFIDENTIAL