

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:
Cardiovascular Assoc. of the Virgin Islands, PLLC
4100 Sion Farm Shopping Center, Suite 5
Christiansted, St. Croix, Virgin Islands 00820

REPORT NUMBER(S) 2014-001

2. NRC/REGIONAL OFFICE

U.S. Nuclear Regulatory Commission
Region I, 2100 Renaissance Blvd, Suite 100
King of Prussia, Pennsylvania 19406-2713

3. DOCKET NUMBER(S)
03037806

4. LICENSE NUMBER(S)
55-31325-01

5. DATE(S) OF INSPECTION
June 23 - Sept 18, 2014

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

Non-cited violation(s) were discussed involving the following requirement(s) and corrective action(s):

- 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with the NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11. (Violations and Corrective Actions)

10 CFR 35.67(g) requires, in part, that a licensee in possession of sealed sources shall conduct a semi-annual physical inventory of all such sources in its possession. The licensee shall retain each inventory record in accordance with 35.2067(b).

Contrary to the above, the licensee did not conduct a semi-annual physical inventory of its sealed sources or retain the inventory record. Specifically, from May 2013 through June 2014, the licensee did not conduct a semi-annual physical inventory of its sealed sources and did not maintain a record of the inventory.

This is a Severity Level IV violation. (Section 6.7)

The licensee committed to placing the sealed sources in storage until work resumes, performing and documenting semi-annual physical inventories of the sealed sources and maintaining records of the inventories as required by 35.2067(b).

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE	Kendall M. Griffith		9/18/14
NRC INSPECTOR	Robin Elliott		9/18/14
BRANCH CHIEF	James P. Dwyer		9/18/14

SUNSI Review Completed By: / RA / RElliott

Public Non-Sensitive