

**From:** [Zarger, Lee A.](#)  
**To:** [Seeley, Shawn](#)  
**Cc:** [Bargellini, Joseph](#)  
**Subject:** RE: New Milford Hospital License renewal MC583901  
**Date:** Thursday, October 30, 2014 10:09:08 AM  
**Attachments:** [Information requested NRC Oct 30.docx](#)

---

Hello Shawn- please see the attached file with answers to your questions.

Lee Anne

---

**Lee Anne Zarger M.S.D.A.B.R.**  
*Chief Medical Physicist*  
*New Milford Hospital*  
*860-210-5022*

---

**From:** Seeley, Shawn [mailto:Shawn.Seeley@nrc.gov]  
**Sent:** Thursday, October 30, 2014 8:34 AM  
**To:** Zarger, Lee A.  
**Cc:** Bargellini, Joseph  
**Subject:** RE: New Milford Hospital License renewal MC583901

Lee Anne,

We have reviewed all the documentation you submitted and have a couple additional questions.

- 1: The shielding diagrams and calculations refer to the planned wall thickness. Please confirm if the planned thickness was actually installed;
- 2: Where are the voids and how are they shielded?;
3. Describe the signs and door interlock and how it is tested daily?;
4. What is your acceptance criteria for the clock date and time check for the correct decay rates for the treatment times?

We look forward to your response. We would like to have this information as quickly as possible as we need to issue this action by the end of next week. If this is going to be an issue, please let me know ASAP. Thank you.

Shawn W. Seeley, Health Physicist  
USNRC-RI, Division of Nuclear Materials Safety  
2100 Renaissance Boulevard, Suite 100  
King of Prussia, PA 19406-2713

Office: 610-337-5102 Fax: 610-337-5269

---

**From:** Zarger, Lee A. [mailto:[Lee.Zarger@wchn.org](mailto:Lee.Zarger@wchn.org)]

**Sent:** Friday, October 03, 2014 2:31 PM  
**To:** Seeley, Shawn  
**Cc:** Bargellini, Joseph  
**Subject:** New Milford Hospital License renewal Information requested

Sean- this is the remainder of the files.  
Please let me know if you need anything else.  
Lee Anne

---

**Lee Anne Zarger M.S.D.A.B.R.**  
*Chief Medical Physicist*  
*New Milford Hospital*  
*860-210-5022*



This transmittal is intended for a particular addressee(s). If it is not clear that you are the intended recipient, you are hereby notified that you have received this transmittal in error; any review, copying or distribution or dissemination is strictly prohibited. If you suspect that you have received this transmittal in error, please notify Western Connecticut Health Network immediately by email reply to the sender, and delete the transmittal and any attachments.

READER BEWARE: Internet e-mail is inherently insecure and occasionally unreliable. Please contact the sender if you wish to arrange for secure communication or to verify the contents of this message.