

The
Cardiology
Group, P.C.

Leonard Grauer, M.D.
Brian C. Swirsky, M.D.
Ron Nudel, M.D.
Jeremy I. Nadelmann, M.D.
Jonathan Brier, M.D.

60 Temple Street
Sixth Floor
New Haven, CT 06510
Telephone (203) 773-3055
FAX (203) 773-9111

06-30162-01

030-33588

March 27, 1995

Ms. Joan Stambaugh
U.S. Nuclear Regulatory Commission
Region I Office
Mail Control #~~12198~~ 121357
475 Allendale Road
King of Prussia, PA 19406

Dear Ms. Stambaugh:


Here are the final surveys and wipe tests of the level 8 scanning area that you requested during the phone conversation with Jonathan Brier, M.D. on March 27, 1995:

All areas in Scanning Room .1 mR/hr.
All areas in Hot Lab - .1 mR/hr.
All areas in Holter and Echo Room - .1 mR/hr.

A wipe test was done of only the counter and sink in Hot Lab area. (The rest of the rooms were completely empty.) The result of the wipe was 186 CTS. and the background was 204 CTS., therefore resulting in 0 DPM.

Thank you very much for your attention. Please feel free to call us with any further questions whatsoever.

Sincerely,



Ann Marie Sigona, RTNM

AMS:jws

OFFICIAL RECORD COPY

121357
ML 10