



Region I Office  
 Division of Nuclear Materials Safety  
 2100 Renaissance Boulevard  
 Suite 100  
 King of Prussia, PA 19406-2713  
 (610)337-5000

## Telephone Conversation Record

Date: December 9, 2014  
 License No.: 06-01060-01  
 Docket No.(no hyphens): 03001247  
 Mail Control/Report No.: 585282  
 Licensee Name: Bridgeport Hospital  
 Participant(s) Name/Title: Michael Bohan - RSO  
 Work Telephone No.: 203-688-2950  
 Business Cellphone No.:  
 NRC Representative Name/Title: T. Weidner/ Sr. HP

**Subject:** RAI concerning PET facility (585282)

**Discussion:** In the letter dated 11/12/2014, Bridgeport Hospital requested the addition of a temporary PET/CT facility to be located at the Medical Office Building currently listed as a LOU on the Bridgeport Hospital license. Additional information about the facility equipment and the occupancy of the telephone closet was received in an e-mail dated 12/3/14. I requested that M. Bohan provide the equipment and phone closet information in a letter signed by Bridgeport Hospital management. In addition, I asked M. Bohan to explain why the shielding in areas requiring greater than 1/16" of lead was going to be installed at 2' from the floor to 6' from the floor only. M. Bohan explained that the installed shielding for the CT portion is 1/16". The lead for the PET shielding is in addition to the 1/16" of lead for the CT. For instance, the shielding for the inner corridor is 1/16" (floor to 7') + 3/4" (2' from floor to 6' from floor).

**Action Required:** Complete the action

SUNSI REVIEW		
<b>Document Availability:</b>	<input checked="" type="checkbox"/> Public	or <input type="checkbox"/> Non-Public
<b>Document Sensitivity:</b> (select "1" value to the right)	<input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive – Proprietary <input type="checkbox"/> Sensitive – Privacy Act (includes Personally Identifiable Information (PII)) <input type="checkbox"/> Sensitive – Internal, Periodic Review required (All Other Sensitive Internal Info.) <input type="checkbox"/> Sensitive – Security-Related-Periodic Review Required	<input type="checkbox"/> MD 3.4 Non-Public B.1 (Non-Sensitive) <input type="checkbox"/> MD 3.4 Non-Public A.3 (Sensitive Security Related) <input type="checkbox"/> MD 3.4 Non-Public A.4 (Sensitive Proprietary) <input type="checkbox"/> MD 3.4 Non-Public A.5 (Sensitive-Privacy Act) <input type="checkbox"/> MD 3.4 Non-Public A.6 (Sensitive-Federal, State, etc) <input type="checkbox"/> MD 3.4 Non-Public A.7 ( Internal)
SUNSI Review Completed by: TLWeidner		

: If applicable, add OFFICIAL USE ONLY – SECURITY RELATED INFORMATION banners and change SUNSI review to A.3.