

CONVERSATION RECORD
(time) (date)

TIME | DATE

10/1/14

VISIT

CONFERENCE X

TELEPHONE X

INCOMING
OUTGOING

NAME OF PERSON(S) CONTACTED OR IN CONTACT

ORGANIZATION (OFFICE, DEPT. ETC.)

TELEPHONE NO.

April Chance, Sr. Manager of Radiation Protection/Environment,
Health and Safety, PETNET

865-218-6355

SUBJECT

Follow-up call to check on status of licensee's response to telephone conversation record dated 9/22/14.

SUMMARY

The licensee's response is due within 15 days of 9/22, and I was calling to check on the status of their response. Ms. Chance stated that one of her staff members was working on the response and she needed to check with him on the status.

I informed Ms. Chance that our expectation is that the licensee will commit to performing manual calculations to determine release activity when daily releases exceed 1200 microcuries since these appear to be more reliable than the software generated calculations. Nevertheless, our plan will be to issue a license condition with this requirement if the licensee does not make this commitment. She indicated that they wanted to verify with staff on site in St. Louis that performing the calculations is feasible. She also stated that most likely it should not be a problem because production has decrease quite a bit in St. Louis and she does not expect that they will reach that level of release very often, if at all.

In addition, I told Ms. Chance it is our understanding that the licensee calibrates the LabImpex system in accordance with the manufacturer's recommended procedure, and that we expect confirmation of this in writing, or we will add a license condition requiring that the licensee calibrate the system per the manufacturer's procedure.

The response is due 15 days from 9/22, which is 10/7/14


NAME OF PERSON DOCUMENTING CONVERSATION

SIGNATURE

DATE

Kevin Null

10/1/14



ACTION TAKEN

SIGNATURE

TITLE

DATE