

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

<p>1. LICENSEE/LOCATION INSPECTED:</p> <p>Digirad Imaging Solutions, Inc. 3825 Paces Walk SE Atlanta, GA 30339 Novi, Michigan clinic</p> <p>REPORT NUMBER(S) 14-001</p>	<p>2. NRC/REGIONAL OFFICE</p> <p>Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352</p>	
<p>3. DOCKET NUMBER(S)</p> <p>030-38357</p>	<p>4. LICENSE NUMBER(S)</p> <p>10-29283-02</p>	<p>5. DATE(S) OF INSPECTION</p> <p>December 16, 2014</p>

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:



- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

Non-cited violation(s) were discussed involving the following requirement(s):

- 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
(Violations and Corrective Actions);

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE	DATE
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Deborah A. Piskura		12/16/14
BRANCH CHIEF	Aaron T. McCraw, Chief, MIB		12/22/14

Docket File Information

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: Digirad Imaging Solutions, Inc. 3825 Paces Walk SE Atlanta, GA 30339 Novi, Michigan clinic REPORT NUMBER(S) 14-001	2. NRC/REGIONAL OFFICE Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352
---	---

3. DOCKET NUMBER(S) 030-38357	4. LICENSE NUMBER(S) 10-29283-02	5. DATE(S) OF INSPECTION December 16, 2014
--------------------------------------	---	---

6. INSPECTION PROCEDURES USED 87130	7. INSPECTION FOCUS AREAS 03.01- 03.07
--	---

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S) 02220	2. PRIORITY 3	3. LICENSEE CONTACT Zubair Abbasi, M.S., RSO	4. TELEPHONE NUMBER (770) 435-9186
---------------------------------	----------------------	---	---

Main Office Inspection Next Inspection Date: no change

Field Office Inspection 24301 Catherine Industrial Dr, Novi, MI

Temporary Job Site Inspection

PROGRAM SCOPE

This was routine inspection of a mobile diagnostic nuclear medicine service authorized to use byproduct material in Sections 35.100 and 35.200 at temporary job sites. The licensee received unit doses at its Novi base hot lab. The licensee performed receipt surveys and repackaged dosages for transport in its company vehicles to client sites. The Novi facility did not possess a dose calibrator and relied on the pharmacy's assay with decay correction. The licensee utilized a small customized solid-state camera for nuclear cardiac stress and rest exams. The Novi site was staffed with one part-time nuclear medicine technologist who routinely conducted an average of 9-10 patient procedures on days when service was provided. The volume of cases has declined and licensee performed studies 2-3 days per week.

This inspection consisted of interviews with licensee personnel, a review of select records, a tour of the nuclear medicine department, and independent measurements. This inspection included one of four facilities authorized on the NRC license. In accordance with MC 2800, one additional site must be inspected by Region III prior to October 2015.

No violations of NRC requirements were identified during this inspection.