

From: [Gallagher, Robert](mailto:Gallagher.Robert)
To: jbean@gccheart.com
Cc: mrosenberg@gccheart.com
Subject: Request for Additional Information - Quality Cardiovascular Care, LLC Renewal Control No. 584750
Date: Tuesday, December 30, 2014 3:32:00 PM
Attachments: [image001.png](#)

PLEASE CONFIRM RECEIPT OF THIS REQUEST FOR ADDITIONAL INFORMATION BY RETURN EMAIL

Dear Mr. Bean,

This is in reference to the application dated September 3, 2014, requesting to renew U.S. Nuclear Regulatory Commission License No. 06-30966-01. In order to continue our review, we need the following additional information:

1. Please confirm you wish the removal of Lily Singhaviranon, M.D. from License No. 06-30966-01.
2. Please provide a diagram to scale (scale used included) that describes the facilities and identifies activities conducted in all contiguous areas surrounding the area(s) of use. The following information should be included:
 - a. Location, room numbers, and principle use of each room or area where byproduct material is prepared, used or stored; and
 - b. Location, room numbers, and principle use of each adjacent room (e.g. office, file, toilet, closet, hallway), including areas above, beside, and below and indicating whether the room is a restricted or unrestricted area as defined in 10 CFR 20.1003.
3. Please provide a statement that: "Either we will perform a prospective evaluation demonstrating that unmonitored individuals are not likely to receive, in 1 year, a radiation dose in excess of 10% of the allowable limits in 10 CFR Part 20 or we will provide dosimetry that meets the requirements listed under 'Criteria' in NUREG-1556, Vol. 9, Rev. 1, 'Consolidated Guidance About Materials Licenses: Program-Specific Guidance About Medical Use Licenses.' "
4. Please provide a statement that: "We have developed and will implement and maintain written procedures for safe use of unsealed byproduct material that meet the requirements of 10 CFR 20.1101 and 10 CFR 20.1301."

Current NRC regulations and guidance are included on the NRC's website at www.nrc.gov; select **Nuclear Materials; Med, Ind, & Academic Uses**; then **Licensee Toolkits**, see our toolkit index page. You may also obtain these documents by contacting the Government Printing Office (GPO) toll-free at 1-866-512-1800. The GPO is open from 8:00 a.m. to 5:30 p.m. EST, Monday through Friday (except Federal Holidays).

We will continue our review upon receipt of this information. Please reply to my attention at the Region I Office and refer to Mail Control No. 584750. If you have any questions regarding this deficiency letter, please call me at the number listed below.

Sincerely,

Robert L. Gallagher
Health Physicist

U.S. NRC, Region I
2100 Renaissance Blvd.
King of Prussia, PA 19406
(610) 337-5182 office
(610) 337-5269 fax

NRC Seal

