



V.I. Quality Control Services, LLC

QA/QC Consultants - Engineers - Technicians
P.O.Box 502911 St.Thomas, Virgin Islands 00805
Phone: 340.344.4717 Fax:340.775.2016
viqcservices@gmail.com

We test the Quality. You have the Control.



January 7, 2015

United States Nuclear Regulatory Commission
Att: Document Control Desk
Blake Welling, Chief
Commercial, Industrial, R&D and Academic
Branch
Division of Nuclear Materials Safety

Re: Reply to Notice of Violation

Dear Mr. Welling:

VI Quality Control Services, LLC is in receipt of your letter dated November 10, 2014. Sorry for the delay in responding to your concerns.

After review and further evaluations, we have made the necessary steps in correcting the violation issued. Please refer to the leak test that was already sent to Mr. Todd Jackson back in November 4, 2014.

The reason this violation occurred is that we were not aware that we could have performed the test ourselves and send the swabs away for results. It was our understanding that the gauge had to be transported for testing to a certified laboratory.

We have now in our position a leak test kit and have determined to test our equipment once a year from here on out.

Should there be any questions, please feel free to contact me via email jking@jubking.com or telephone 340.690.6407 at your earliest convenience.

Respectfully,

Julio U.B. King, RSO
VIQC Services, LLC