



January 8, 2014

UNITED STATES NUCLEAR REGULATORY COMMISSION  
Region III, Materials Licensing Section  
2443 Warrenville Road, Suite 210  
Lisle, IL 60532-4352

**Re: Additional Information to Reference No.: CN 585182  
License No. 21-12829-01, Metro Health Hospital.**

Dear Mr. Tran:

In addition to the information submitted in our initial request for authorization of the use of Y-90 Sir-Sphere microspheres we commit to the following:

When the Y-90 microspheres are placed in vials, syringes, or radiation shields that are not labeled by the manufacturer, the licensee shall:

- 1) Label vials and vial radiation shields with radionuclide and form (e.g., Y-90 microspheres).
- 2) Label syringes and syringe radiation shields with the radionuclide, form, and therapeutic procedure (e.g., Y-90 microspheres, brachytherapy).

If you have any questions regarding this matter, please contact our Physics Consultant, Dawn Edwards, at 734-662-3197 or [dedwards@mpcphysics.com](mailto:dedwards@mpcphysics.com).

Sincerely,

A handwritten signature in black ink, appearing to read "Dawn Edwards".

**Executive Management**

RECEIVED JAN 13 2015



FAX COVER SHEET

2

TO: Mr. Tran

FROM: Panopoulos Imaging Center  
Metro Health Hospital  
Radiology Department  
5900 Byron Center Avenue SW  
Wyoming, MI 49519

CC: (630) 515-1078

Phone 616 252-7159  
Fax Phone 616 252-6246

REMARKS:  Urgent  For your review  Reply ASAP  Please Comment

Please review the attached document for important information.  
Thank you

Mr. Tran,  
Attached is additional  
document / information on  
metro letter head as requested.  
Jezza Kaso

**CONFIDENTIAL INFORMATION**  
If you receive this information in error, please call  
the sending phone number immediately!  
Thank you.