

CONVERSATION RECORD

TIME DATE 1/22/15

0 VISIT

0 CONFERENCE

X TELEPHONE

0 INCOMING
X OUTGOING

NAME OF PERSON(S) CONTACTED OR IN CONTACT
Mark Driscoll, RSO

ORGANIZATION (OFFICE, DEPT. ETC.)
University of Michigan

TELEPHONE NO.
734-647-2251

SUBJECT

Condition 24.E. of Amendment No. 101 of License No. 21-00215-04 (Docket No.: 03001988)

SUMMARY

Based on follow-up of a violation that was identified during a recent inspection conducted at another licensee's facilities, the inspector informed the RSO that Condition 24.E. of Amendment No. 101 of License No. 21-00215-04 states, "Immediately stop the use of the irradiator and notify the NRC, Region III, by telephone if abnormal levels of radiation or any malfunction of the irradiator is detected." That verbiage was superseded by new verbiage on Page R-6 of NUREG-1556, Vol. 5. (i.e.; "If abnormal radiation levels or any malfunctions of the irradiator are detected at any time, stop using the irradiator, restrict access to the area housing the irradiator, immediately notify the Radiation Safety Officer, and submit all reports required under 10 CFR Parts 20, 21 or 30." The RSO stated that he would review this issue and determine whether or not he would submit a license amendment request to change Condition 24.E. of Amendment No. 101 of License No. 21-00215-04.

ACTION REQUIRED

NAME OF PERSON DOCUMENTING CONVERSATION
Bob Gattone

SIGNATURE

DATE

Robert D. Gattone Jr.

1/22/15

ACTION TAKEN

SIGNATURE

TITLE

DATE