

Region I Office Division of Nuclear Materials Safety 2100 Renaissance Boulevard Suite 100 King of Prussia, PA 19406-2713 (610)337-5000

Telephone Conversation Record

Date: License No.: Docket No.(no hyphens): Mail Control/Report No.: Licensee Name: Participant(s) Name/Title: Work Telephone No.: Business Cellphone No.: NRC Representative Name/Title: 1.23.15 06-02388-01 03001250 585522 Hospital of Central Connecticut Eric Lyons, Manager 860-224-5011

T. Weidner/Sr. HealthPhysicist

Subject: PET location of use

Discussion: In the letter dated December 8, 2014, HOCC requested the addition of a PET facility at 201 N. Mountain Road, Plainville, CT. On 1.23.15 I spoke to Eric Lyons for clarification of the project. According to E. Lyons, the PET facility at the New Britain LOU is being decommissioned and the equipment used there is being transferred to the Mountain Road LOU. This includes the camera and all hot lab related equipment (L-block, syringe shields, shielded well counter, etc.). E. Lyons will provide the following additional information for the PET facility:

- 1. Readable facility diagrams
- 2. Description of the areas above and below the use areas, and
- 3. Shielding design for the PET areas of use

Action Required: Wait for additional information

SUNSI REVIEW				
Document Availability:	ailability: 🗴 Public or 🔄 Non-Public			
Document Sensitivity:				
(select "1" value to the right)	х	Non-Sensitive		MD 3.4 Non-Public B.1 (Non-Sensitive)
		Non-Sensitive Copyright		MD 3.4 Non-Public A.3 (Sensitive Security Related
		Sensitive – Proprietary		MD 3.4 Non-Public A.4 (Sensitive Proprietary
		Sensitive – Privacy Act (includes Personally Identifiable Information (PII))		MD 3.4 Non-Public A.5 (Sensitive-Privacy Act)
		Sensitive – Internal, Periodic Review required (All Other Sensitive Internal Info.)		MD 3.4 Non-Public A.6 (Sensitive-Federal, State, etc)
		Sensitive – Security-Related-Periodic Review Required		MD 3.4 Non-Public A.7 (Internal)
SUNSI Review Completed by:	eted by: TLWeidner			

: If applicable, add OFFICIAL USE ONLY – SECURITY RELATED INFORMATION banners and change SUNSI review to A.3.