



CONVERSATION RECORD

01/12/2015

NAME OF PERSON(S) CONTACTED OR IN CONTACT WITH YOU Michelle Kritzman		DATE OF CONTACT 01/12/2015	TYPE OF CONVERSATION <input type="checkbox"/> E-MAIL <input checked="" type="checkbox"/> TELEPHONE <input checked="" type="checkbox"/> INCOMING <input type="checkbox"/> OUTGOING
E-MAIL ADDRESS mkritzman@mpcphysics.com		TELEPHONE NUMBER (734) 662-3197	

ORGANIZATION Medical Nuclear Physicist Medical Physics Consultants	DOCKET NUMBER(S) 030-02027
--	-------------------------------

LICENSE NUMBER(S) 21-03194-01	CONTROL NUMBER(S) 584593
----------------------------------	-----------------------------

SUBJECT
Additional Information Needed for ProMedica Bixby Hospital License Renewal

SUMMARY

Based on our review of your application to renew your NRC license, there are several questions we would like for you to address:

1. On your current license in Condition 10.B, you are authorized to use 10 CFR 35.100 and 35.200 materials at any medical care facility in the U.S. where the NRC has jurisdiction. The authorization for mobile medicine was not requested in the license renewal application.

Please confirm in writing that you whether or not you plan to continue using licensed materials at other medical care facilities. If you will be using licensed materials at other medical care facilities, please provide a description of your program and all the procedures regarding the use of radioactive materials at temporary job sites. Please be sure to include where the radioactive materials will be received, how they will be transported to the temporary job site, emergency procedures and how surveys will be conducted after each administration.

Continue on Page 2

ACTION REQUIRED (IF ANY)

Please submit your response by January 23, 2014, and reference it to my attention as "additional information to control number 584593" to facilitate proper handling in our office. Your response must be currently dated and signed by a senior management representative. If you have any questions or require clarification of any of the information stated above, please do not hesitate to contact me at 630-829-9607

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter and its enclosure will be available electronically for public inspection in the NRC Public Document Room or from the NRC's Agencywide Documents Access and Management System (ADAMS), accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>.

Continue on Page 3

NAME OF PERSON DOCUMENTING CONVERSATION
Jennifer L. Bishop

SIGNATURE

CONVERSATION RECORD (continued)

SUMMARY: (Continued from page 1)

2. In your application you have a Gadolinium-153 source listed, however you do not indicated what should be the authorized use for this source. Your current license has that the source as 'possession and storage only incident to disposal.'

Please confirm in writing what the authorized use for the source should be.

3. While reviewing the application, we noticed that there appear to be several Authorized Users listed on you application, and in your most current NRC license, more than once. In addition, you requested to remove several Authorized Users from the license and some of these individuals appear to be on the license twice.

To ensure the license is correct, please review and provide an updated list of all the Authorized Users you that should be listed on your license.

4. On pages 6 and 7 of your application, you include facility diagrams. At the bottom Page 6 you indicate what is directly above and below the area in the drawings, but the drawings are not clear. On Page 7 you state that there is a hot lab, however the location of the hot lab is not clearly indicated.

Please either confirm the drawings are correct or provide a new drawings of the areas with the correct information.