



**UNITED STATES
NUCLEAR REGULATORY COMMISSION**

REGION III
2443 WARRENVILLE RD. SUITE 210
LISLE, IL 60532-4352

October 21, 2014

Anthony A. Bennett, M.D.
Radiation Safety Officer
Complete Health Systems
5084 Villa Linde Parkway
Suite 7
Flint, MI 48532

Dear Dr. Bennett:

We have reviewed your renewal application dated August 6, 2014. In order to complete our review, we will need the following information.

- 1) Please confirm that Mr. Craig VanHouten, CNMT, was authorized by the management to sign the renewal application dated August 6, 2014.
- 2) Please describe the type of use of your mobile medical service (do you perform imaging procedures in the van or in your client facilities?) and a readable diagram of the mobile unit with scale or dimensions.
- 3) Please confirm that you do not use PET isotopes.
- 4) If you plan to perform imaging procedures at 5084 Villa Linde Pkwy, Flint, Michigan, please identify the imaging (camera) room in the diagram.
- 5) Please confirm that byproduct material will only be delivered to the van when it is occupied by the licensee authorized personnel.
- 6) Please provide a statement that the licensee will either perform a prospective evaluation demonstrating that unmonitored individuals are not likely to receive, in one year, a radiation dose in excess of 10% of the allowable limits in 10 CFR Part 20 or the licensee will provide dosimetry that meets the requirements listed under "Criteria" in NUREG-1556, Volume 9, Revision 2, "Consolidated Guidance About Materials Licenses: Program-Specific Guidance About Medical Use Licenses".
- 7) Please provide a statement that the licensee has developed and implemented and maintained written procedures for area surveys in accordance with 10 CFR 20.1101 that meet the requirements of 10 CFR 20.1501 and 10 CFR 35.70.


NUREG-1556, Volume 9, Rev. 2, dated January 2008 provides the program specific guidance about medical use licenses. NUREG-1556, Volume 9 is available on the NRC website at <http://www.nrc.gov/reading-rm/doc-collections/nuregs/staff/sr1556/v9/r2/>.

To continue the review of your request, we request that you submit your response to this letter by November 12, 2014. Your response must be dated and signed by authorized personnel and

referred to Control Number 584552 to facilitate proper mail handling in our office. If you have any questions or require clarification on any of the information stated above, please do not hesitate to contact me at 630-829-9623 or frank.tran@nrc.gov.

In accordance with Title 10 Code of Federal Regulations 2.390 of the NRC's "Rules of Practice," a copy of this letter will be available electronically for public inspection in the NRC Public Document Room or from the NRC's Agencywide Documents Access and Management System (ADAMS), accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>.

Sincerely,



Frank P.D. Tran
Health Physicist
Materials Licensing Branch

License No. 21-32543-01
Docket No. 030-36714

Tran, Frank

From: Tran, Frank
Sent: Tuesday, October 21, 2014 2:01 PM
To: 'v_mccormick@comcast.net'
Subject: Request for additional information re Complete Health Systems
Attachments: Scan001.PDF

Dear Mr. McCormick,

We have reviewed the renewal application for Complete Health Systems, NRC License No. 21-32543-01. We will need additional information to complete our review (see attached).

If you have any questions or need a clarification regarding the request, please contact me at 630-829-9623 or reply to this email.

Thank you,

Frank Tran

-----Original Message-----

From: R3-238-DNMS-B@nrc.gov [mailto:R3-238-DNMS-B@nrc.gov]
Sent: Tuesday, October 21, 2014 3:48 PM
To: Tran, Frank
Subject: Scan from R3-238-DNMS-B

Sent by: Guest [R3-238-DNMS-B@nrc.gov]
Number of Images: 2
Attachment File Type: PDF

Device Name: R3-238-DNMS-B