

Quality Cardiovascular Care
1389 W Main St
Waterbury CT 06708

REC'D 10127 15AM 1106

January 13, 2015

Mr. Robert Gallagher
USNRC Region I
2100 Renaissance Blvd
King of Prussia PA 19406

J3

03036684
06-30966-01


Greetings:

In response to the deficiency letter received December 30, 2014 for control number 584750, we have the following response:

1. We confirm we wish the removal of Lily Singhaviranon, M.D. from License No. 06-30966-01.
2. The information you requested regarding area diagram and shielding survey was included in the renewal application dated September 3, 2014. We do not use any PET agents. The Hot-Lab as indicated on the diagram is the only restricted area. The other areas of the facility are unrestricted.
3. Either we will perform a prospective evaluation demonstrating that unmonitored individuals are not likely to receive, in 1 year, a radiation dose in excess of 10% of the allowable limits in 10 CFR Part 20 or we will provide dosimetry that meets the requirements listed under 'Criteria' in NUREG-1556, Vol. 9, Rev. 1, 'Consolidated Guidance About Materials Licenses: Program-Specific Guidance About Medical Use Licenses.
4. We have developed and will implement and maintain written procedures for safe use of unsealed byproduct material that meet the requirements of 10 CFR 20.1101 and 10 CFR 20.1301.

Please contact us if you require additional information.

Sincerely yours,



Mark Rosenberg, MD

MR/srk

584750

NMSS/RGN1 MATERIALS-002