

**PATRIOT ENGINEERING
and Environmental, Inc.**

Engineering Value for Project Success

February 11th, 2015
United States Nuclear Regulatory Commission
Division of Nuclear Materials Safety
243 Warrenville Road, Suite 210
Lisle, IL 60532

Re: License Number 13-32725-01
Docket Number 03037878
Control Number 584382
Conversation Record Date of Signature February 9, 2015
Items #4

Dear Ms. Casey;

This letter is intended to address item #4 of the referenced conversation record.

Item #4


Per your request as outlined in the referenced conversation record, our current operational address for our Evansville, IN office is 601 East Sycamore Street, Evansville, IN 47113. We have retained storage facilities at 104 South Heidelberg Avenue, Evansville, IN in the event these gauges are not in use.

In addition, and in accordance with our conversation yesterday and to clarify our request from July 15, 2014, Patriot would like to add our office at 601 East Sycamore Street, Evansville, IN 47113 to our license concurrent with the removal of our facility located at 104 South Heidelberg Avenue, Evansville, IN.

6330 East 75th Street, Suite 216, Indianapolis, Indiana 46250
(317) 576-8058 • (317) 576-1965 FAX • www.patrioteng.com

Offices in Indianapolis, Evansville, Fort Wayne, Lafayette, Bloomington and Terre Haute, IN,
Louisville, KY, Dayton / Cincinnati, OH, Nashville, TN, Carmel, IL, and New Orleans, LA

RECEIVED FEB 12 2015




**PATRIOT ENGINEERING
and Environmental, Inc.**

Engineering Value for Project Success

Included with this letter, please find attached our most recent leak test results for each gauge we intend to move to the 601 East Sycamore Street, Evansville, IN address.

Again, we appreciate your assistance in resolving this matter. If you have any questions, please do not hesitate to contact us.

Respectfully submitted,
Patriot Engineering and Environmental Inc.


James T. Sherer
President


Bryan King
Corporate RSO

6330 East 75th Street, Suite 216, Indianapolis, Indiana 46250
(317) 576-8058 • (317) 576-1965 FAX • www.patrioteng.com

Offices in Indianapolis, Evansville, Fort Wayne, Lafayette, Bloomington and Terre Haute, IN,
Louisville, KY, Dayton / Cincinnati, OH, Nashville, TN, Carmi, IL, and New Orleans, LA

Seaman Nuclear Corporation

7216 South First Street Oak Creek, WI 53184 USA
Tel 414-782-5100 Fax 414-782-5100
info@seamannuclear.com

Leak Test Certificate

September 15, 2014

A leak test has been performed on meter, serial number 7830, a model C-76-A, containing the radionuclide Ra 226

Owned By: Patriot Eng. & Environmental Inc.
8330 E. 78th St., Suite 218
Indianapolis, IN 46260

Date Sample Collected: 9/8/14
Collected By: Adam Bennett
Date Sample Analyzed: 9/16/14
Analyzed By: K. Krystek

Most regulatory agencies consider a source to be leaking if a leak test reveals the presence of more than 0.005 microcurie of removable contamination.

Analysis found contamination of less than 0.005 microcurie.

Analysis authorized by Wisconsin license 079-1267-01.

LEAK TEST DUE: 3/8/15

Scott C. Seaman
Radiation Safety Officer
Seaman Nuclear Corporation

Seaman Nuclear Corporation

7315 South First Street Oak Creek, WI 53154 USA
Tel 414-762-5100 Fax 414-762-5106
info@seamannuclear.com

Leak Test Certificate

September 8, 2014

A leak test has been performed on meter, serial number 7522, a model C-75-A, containing the radionuclide Ra 226

Owned By: Patriot Eng. & Environmental Inc.
6330 E. 75th St., Suite 216
Indianapolis, IN 46250


Date Sample Collected: 9/3/14
Collected By: Brian Crase
Date Sample Analyzed: 9/8/14
Analyzed By: K. Krysiak

Most regulatory agencies consider a source to be leaking if a leak test reveals the presence of more than 0.005 microcurie of removable contamination.

Analysis found contamination of less than 0.005 microcurie.

Analysis authorized by Wisconsin license 079-1257-01.

LEAK TEST DUE: 3/3/15



Scott C. Seaman
Radiation Safety Officer
Seaman Nuclear Corporation

S/N: 7522
Leak Tested: 9/3/14
Due: 3/3/15
Seaman Nuclear Corporation
Oak Creek, WI (414) 762-5100

Seaman Nuclear Corporation

7315 South First Street Oak Creek, WI 53154 USA
Tel 414-762-5100 Fax 414-762-5100
Info@seamannuclear.com

Leak Test Certificate

September 8, 2014

A leak test has been performed on meter, serial number 7478, a model C-75-A, containing the radionuclide Ra 226

Owned By: Patriot Eng. & Environmental Inc.
6330 E. 75th St., Suite 216
Indianapolis, IN 46250


Date Sample Collected: 9/3/14
Collected By: Brian Crase
Date Sample Analyzed: 9/8/14
Analyzed By: K. Krysiak

Most regulatory agencies consider a source to be leaking if a leak test reveals the presence of more than 0.005 microcurie of removable contamination.

Analysis found contamination of less than 0.005 microcurie.

Analysis authorized by Wisconsin license 079-1257-01.

LEAK TEST DUE: 3/3/15



Scott C. Seaman
Radiation Safety Officer
Seaman Nuclear Corporation

S/N: 7478
Leak Tested: 9/8/14
Due: 3/3/15
Seaman Nuclear Corporation
Oak Creek, WI (414) 762-5100

Seaman Nuclear Corporation

7315 South First Street Oak Creek, WI 53154 USA

Tel 414-762-5100 Fax 414-762-5106

info@seamannuclear.com

Leak Test Certificate

September 8, 2014

A leak test has been performed on meter, serial number **7524**, a model C-75-A, containing the radionuclide Ra 226

Owned By: Patriot Eng. & Environmental Inc.
6330 E. 75th St., Suite 216
Indianapolis, IN 46250


Date Sample Collected: 9/3/14
Collected By: Brian Crase
Date Sample Analyzed: 9/8/14
Analyzed By: K. Krysiak

Most regulatory agencies consider a source to be leaking if a leak test reveals the presence of more than 0.005 microcurie of removable contamination.

Analysis found contamination of less than 0.005 microcurie.

Analysis authorized by Wisconsin license 079-1257-01.

LEAK TEST DUE: 3/3/15



Scott C. Seaman
Radiation Safety Officer
Seaman Nuclear Corporation

S/N: 7524
Leak Tested: 9/3/14
Due: 3/3/15
Seaman Nuclear Corporation
Oak Creek, WI (414) 762-5100

Seaman Nuclear Corporation

7315 South First Street Oak Creek, WI 53154 USA
Tel 414-782-5100 Fax 414-782-5108
Info@seamannuclear.com

Leak Test Certificate

September 8, 2014

A leak test has been performed on meter, serial number 7502, a model C-75-A, containing the radionuclide Ra 226

Owned By: Patriot Eng. & Environmental Inc.
6330 E. 75th St., Suite 216
Indianapolis, IN 46250

Date Sample Collected: 9/3/14
Collected By: Brian Crase
Date Sample Analyzed: 9/8/14
Analyzed By: K. Kryslak

Most regulatory agencies consider a source to be leaking if a leak test reveals the presence of more than 0.005 microcurie of removable contamination.

Analysis found contamination of less than 0.005 microcurie.

Analysis authorized by Wisconsin license 079-1257-01.

LEAK TEST DUE: 3/3/15



Scott C. Seaman
Radiation Safety Officer
Seaman Nuclear Corporation

S/N: 7502
Leak Tested: 9/3/14
Due: 3/3/15
Seaman Nuclear Corporation
Oak Creek, WI (414) 782-5100

Seaman Nuclear Corporation

7315 South First Street Oak Creek, WI 53154 USA

Tel 414-762-5100 Fax 414-762-5106

info@seamannuclear.com

Leak Test Certificate

September 8, 2014

A leak test has been performed on meter, serial number 7512, a model C-75-A, containing the radionuclide Ra 226

Owned By: Patriot Eng. & Environmental Inc.
6330 E. 75th St., Suite 216
Indianapolis, IN 46250


Date Sample Collected: 9/3/14
Collected By: Brian Crase
Date Sample Analyzed: 9/8/14
Analyzed By: K. Krysiak

Most regulatory agencies consider a source to be leaking if a leak test reveals the presence of more than 0.005 microcurie of removable contamination.


Analysis found contamination of less than 0.005 microcurie.

Analysis authorized by Wisconsin license 079-1257-01.

LEAK TEST DUE: 3/3/15



Scott C. Seaman
Radiation Safety Officer
Seaman Nuclear Corporation

S/N: 7512
Leak Tested: 9/3/14
Due: 3/3/15
 Seaman Nuclear Corporation
Oak Creek, WI (414) 762-5100

Seaman Nuclear Corporation

7315 South First Street Oak Creek, WI 53154 USA
Tel 414-762-5100 Fax 414-762-5106
info@seamannuclear.com

Leak Test Certificate

September 8, 2014

A leak test has been performed on meter, serial number **1206**, a model Humboldt, containing the radionuclide Cs 137, Am 241

Owned By: Patriot Eng. & Environmental Inc.
400 Production Court
Louisville, KY 40299


Date Sample Collected: 9/3/14
Collected By: Brian Crase
Date Sample Analyzed: 9/8/14
Analyzed By: K. Krysiak

Most regulatory agencies consider a source to be leaking if a leak test reveals the presence of more than 0.005 microcurie of removable contamination.

Analysis found contamination of less than 0.005 microcurie.

Analysis authorized by Wisconsin license 079-1257-01.

LEAK TEST DUE: 3/3/15



Scott C. Seaman
Radiation Safety Officer
Seaman Nuclear Corporation

S/N: 1206
Leak Tested: 9/3/14
Due: 3/3/15
Seaman Nuclear Corporation
Oak Creek, WI (414) 762-5100

Seaman Nuclear Corporation

7315 South First Street Oak Creek, WI 53154 USA
Tel 414-782-5100 Fax 414-782-5108
Info@seamannuclear.com

Leak Test Certificate

September 16, 2014

A leak test has been performed on meter, serial number 7586, a model C-75-A, containing the radionuclide Ra 226

Owned By: Patriot Eng. & Environmental Inc.
1359 North Aberdeen Ave.
Terra Haute, IN 47804


Date Sample Collected: 9/3/14
Collected By: Brian Crase
Date Sample Analyzed: 9/8/14
Analyzed By: K. Krysiak

Most regulatory agencies consider a source to be leaking if a leak test reveals the presence of more than 0.005 microcurie of removable contamination.

Analysis found contamination of less than 0.005 microcurie.

Analysis authorized by Wisconsin license 079-1257-01.

LEAK TEST DUE: 3/3/15



Scott C. Seaman
Radiation Safety Officer
Seaman Nuclear Corporation

Seaman Nuclear Corporation

7315 South First Street Oak Creek, WI 53154 USA

Tel 414-762-5100 Fax 414-762-5106

Info@seamannuclear.com

Leak Test Certificate

September 8, 2014

A leak test has been performed on meter, serial number 4631, a model Humboldt, containing the radionuclide Cs 137, Am 241

Owned By: Patriot Eng. & Environmental Inc.
400 Production Court
Louisville, KY 40299


Date Sample Collected: 9/5/14
Collected By: N. Durkee
Date Sample Analyzed: 9/8/14
Analyzed By: K. Krysiak

Most regulatory agencies consider a source to be leaking if a leak test reveals the presence of more than 0.005 microcurie of removable contamination.


Analysis found contamination of less than 0.005 microcurie.

Analysis authorized by Wisconsin license 079-1257-01.

LEAK TEST DUE: 3/5/15



Scott C. Seaman
Radiation Safety Officer
Seaman Nuclear Corporation

S/N: 4631
Leak Tested: 9/5/14
Due: 3/15/15
 Seaman Nuclear Corporation
Oak Creek, WI (414) 762-5100

Seaman Nuclear Corporation

7315 South First Street Oak Creek, WI 53154 USA
Tel 414-782-5100 Fax 414-782-5108
Info@seamannuclear.com

Leak Test Certificate

September 8, 2014

A leak test has been performed on meter, serial number 20494, a model Other, containing the radionuclide None

Owned By: Patriot Eng. & Environmental Inc.
400 Production Court
Louisville, KY 40299


Date Sample Collected: 9/5/14
Collected By: N. Durkee
Date Sample Analyzed: 9/8/14
Analyzed By: K. Krysiak

Most regulatory agencies consider a source to be leaking if a leak test reveals the presence of more than 0.005 microcurie of removable contamination.


Analysis found contamination of less than 0.005 microcurie.

Analysis authorized by Wisconsin license 079-1257-01.

LEAK TEST DUE: 3/5/15



Scott C. Seaman
Radiation Safety Officer
Seaman Nuclear Corporation

S/N: 20494
Leak Tested: 9/5/14
Due: 3/5/15
 Seaman Nuclear Corporation
Oak Creek, WI (414) 782-5100

Seaman Nuclear Corporation

7315 South First Street Oak Creek, WI 53154 USA
Tel 414-762-5100 Fax 414-762-5106
Info@seamannuclear.com

Leak Test Certificate

September 9, 2014

A leak test has been performed on meter, serial number **18538**, a model Troxler, containing the radionuclide Am 241

Owned By: Patriot Eng. & Environmental Inc.
104 S. Heidelberg Ave., Suite A
Evansville, IN 47713

Date Sample Collected: 9/3/14
Collected By: Brian Crase
Date Sample Analyzed: 9/8/14
Analyzed By: K. Krysiak


9/9/14
K. Krysiak

Most regulatory agencies consider a source to be leaking if a leak test reveals the presence of more than 0.005 microcurie of removable contamination.

Analysis found contamination of less than 0.005 microcurie.

Analysis authorized by Wisconsin license 079-1257-01.

LEAK TEST DUE: 3/3/15



Scott C. Seaman
Radiation Safety Officer
Seaman Nuclear Corporation

S/N: 18538
Leak Tested: 9/8/14
Due: 3/3/15
Seaman Nuclear Corporation
Oak Creek, WI (414) 762-5100



6330 East 75th Street, Suite 216
Indianapolis, IN 46250-2700
317-576-8058
Fax: 317-570-7182

FAX TRANSMISSION

To: COLLEEN CASEY

Date: 2/11/2015

Company: Nuclear Regulatory Commission

Fax No.: 630-515-1078

From: Bryan King

Number of Pages: 14
(Including Cover Page)

Remarks: