From: Seeley, Shawn
To: Ivette-777@hotmail.com
Cc: Elliott, Robin
Subject: Request for Additional Information MC585977
Date: Friday, March 06, 2015 8:53:00 AM

Dear Dr. Quiñones:

Please acknowledge receipt of this email.

This is in reference to your application dated January 14, 2015, requesting to renew Nuclear Regulatory Commission License No. 52-23038-01. In order to continue our review, we need the following additional information:

1. In order to facilitate inspection and communication, please provide GPS coordinates for your facility if they are available and a business facsimile number.

2. Your current license lists two addresses of use: 111 Font Martelo Avenue and Hospital Font Martelo, Inc., 3 Font Martelo Avenue. Your renewal application lists only the 111 Font Martelo Avenue address. Do you wish to retain your authorization to use your Strontium 90 eye applicator at Hospital Font Martelo, Inc.?

3. Item 9 of your application states that a facility diagram is enclosed; however, we did not receive the diagram. Please provide a detailed version of your facility diagram indicating the locations of use and storage of your Strontium 90 eye applicator. Include location, room numbers, and principal use of each adjacent room (e.g., office, file, toilet, closet, hallway), including areas above, beside, and below therapy treatment rooms. Confirm that public dose limits will be met in unrestricted areas when the source is in storage and use.

4. You indicated that the source is locked in a safe when the authorized user is not present. Please indicate where the key/combination is kept and who has access to it.

Current NRC regulations and guidance are included on the NRC's website at www.nrc.gov; select **Nuclear Materials; Med, Ind, & Academic Uses;** then **Licensee Toolkits, see our toolkit index page.** You may also obtain these documents by contacting the Government Printing Office (GPO) toll-free at 1-866-512-1800. The GPO is open from 8:00 a.m. to 5:30 p.m. EST, Monday through Friday (except Federal holidays).

We will continue our review upon receipt of this information. Please reply to my attention at the Region I Office and refer to Mail Control No. 585977. If you have any technical questions regarding this deficiency letter, please call Shawn Seeley at (610) 337-5102.

In order to continue prompt review of your application, we request that you submit your response to this letter within 30 calendar days from the date of this letter.

Sincerely,

Robin Elliott
Health Physicist
Medical Branch
Division of Nuclear Materials Safety

Shawn W. Seeley, Health Physicist
USNRC-RI, Division of Nuclear Materials Safety
2100 Renaissance Boulevard, Suite 100
King of Prussia, PA 19406-2713

Office: 610-337-5102 Fax: 610-337-5269