



**UNITED STATES
NUCLEAR REGULATORY COMMISSION**
REGION I
2100 RENAISSANCE BOULEVARD, SUITE 100
KING OF PRUSSIA, PA 19406-2713

March 12, 2015

Docket No. 03001274
Control No. 585574

License No. 06-08544-01

Daniel DeBarba
President
Danbury Hospital
24 Hospital Avenue
Danbury, CT 06810

**SUBJECT: DANBURY HOSPITAL, VOIDANCE OF APPLICATION FOR LICENSE
AMENDMENT, CONTROL NO. 585574**

Dear Mr. DeBarba:

This concerns the subject application for an Amendment to your material license to add microsphere use. Because the additional information required to complete the requested action is delayed for some time, we have voided your application. This action is taken without prejudice to the resubmission of your request. In a future request, please refer to your December 5, 2014 letter and provide the following additional information originally requested on February 12, 2014:

1. Confirm that the Radiation Safety Officer attended training on the use of microspheres that occurred on November 14, 2014.
2. Provide a copy of the training memo for the training provided to Dr. Sanghavi by the manufacturer.
3. Document 80 hours of classroom and laboratory training for byproduct material, including Y-90 microspheres, for Dr. Sandhu to include radiation physics and instrumentation, radiation protection, mathematics pertaining to the use and measurement of radioactivity, and radiation biology.
4. Revise Item 10 to remove the references to revised regulations since the intent is to allow changes to the licensee's procedures when the NRC's suggested procedures change and not when regulations change.

Sincerely,

Original signed by Penny Lanzisera

Penny Lanzisera
Senior Health Physicist
Medical Branch
Division of Nuclear Materials Safety

D. DeBarba

2

cc:
Vladimir Monastyrenko, Ph.D., Radiation Safety Officer

DOCUMENT NAME: G:\WordDocs\Current\Lic Cvr Letter\L06-08544-01.585574.doc

SUNSI Review Complete: PLanzisera

After declaring this document "An Official Agency Record" it will be released to the Public.

To receive a copy of this document, indicate in the box: "C" = Copy w/o attach/encl "E" = Copy w/ attach/encl "N" = No copy

OFFICE	DNMS/RI	N					
NAME	PLanzisera/PL						
DATE	03/12/2015						

OFFICIAL RECORD COPY