

U.S. NUCLEAR REGULATORY COMMISSION
TELEPHONE CONVERSATION RECORD

Date: March 26, 2015
1:00 pm

Inspection No. 2015-001 License No. 47-25478-01 Docket No. 030-35150

Name of Licensee: Cardiovascular Consultants, PLLC
Name of Participant(s): Gerry Spears, Office Manager; Amy Griffith, CNMT; Dr. Haffar, RSO/Owner
Shawn Seeley, Health Physicist, RI DNMS
Telephone No. 304-757-4694
Subject: Inspection Exit Meeting and Next Steps
(NOTE: This will be used as the Documents Title in ADAMS)

I called Dr. Haffar's office to conduct a final exit meeting from my February 19, 2015, onsite inspection with continued in-office review through March 26, 2015. I told them that they would be getting a letter and notice of violation with five apparent violations. I mentioned to them that they would have 20 days to provide us a response. The violations were the failure to:

- 1: conduct a six month leak test;
- 2: conduct annual survey instrument calibrations;
- 3: conduct quarterly dose calibrator linearity tests;
- 4: conduct adequate DOT HAZMAT training; and
- 5: conduct an adequate annual audit

They discussed the following corrective actions:

They have entered reminders in their CENTREX nuclear medicine system for when required items are due, will contact their consultant regarding the annual audit deficiencies, and are committed to ensuring safety and compliance with the regulations. They would respond once they received the notification.

I thanked them for their time and ended the call.

Document Availability:

Publicly Available

Non-Publicly Available

Non-Sensitive
 Sensitive – Privacy Act/ PII
 Immediate Release

Non-Sensitive Copyright
 Sensitive – Internal
Normal Release Date:

Sensitive- Proprietary
 Sensitive – Security-Related
Delayed Release Date:

SUNSI Review Completed By: SSeeley / RA /

Document Accession #: _____