



**UNITED STATES
NUCLEAR REGULATORY COMMISSION**

REGION IV
1600 E. LAMAR BLVD.
ARLINGTON, TX 76011-4511

April 20, 2015

Mr. Steve Lytle
Radiation Safety Officer
Spokane Rock Products, Inc.
P.O. Box 3808
Spokane, Washington 99220-3808

SUBJECT: TERMINATION OF NRC RADIOACTIVE MATERIALS LICENSE NO. 46-27773-01

By NRC Form 314 dated January 21, 2015, you informed the U.S. Nuclear Regulatory Commission (NRC) that Spokane Rock Products, Inc., wished to terminate NRC radioactive materials license number 46-27773-01, a license that permitted use of portable gauge devices at temporary job sites. The NRC Form 314 dated January 21, 2015 stated that all radioactive materials had been disposed by transfer to a licensed entity.

The supporting documentation to the NRC Form 314 provided copies of the most recent leak test results. The leak test results showed that there was no evidence of leakage of radioactive material from the sealed sources. Based on the documentation submitted and a review of your historical file, the staff concluded that all licensable radioactive material has been removed from your site. Based on these conclusions, no further remediation or actions with respect to NRC regulated material is required.

Please find enclosed Amendment No. 04 to License No. 46-27773-01 authorizing your license termination. An environmental assessment for this action is not required since this action is categorically excluded under 10 CFR 51.22(c)(20)(ii).

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter and its enclosure will be available electronically for public inspection in the NRC Public Document Room or from the NRC's document system (ADAMS). ADAMS is accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>.

You can contact me at 817-200-1273 if you have questions or require clarification on any of the information stated above.

Sincerely,

/RA/

Gerald Schlapper, CHP, Health Physicist
Nuclear Materials Safety Branch B

Docket: 030-36553
License: 44-27773-01
Control: 585883

Enclosure: As stated