



MISTRAS

Asset Protection Solutions

Services
Division

P: (281) 478-1600 • F: (281) 478-1781
4000 Underwood Rd. La Porte, TX 77571
www.mistrasgroup.com

April 27, 2015

U. S. NRC Region III
2443 Warrenville Road, Suite 210
Lisle, IL 60532-4352

Re: Removing location from RAM License 12-16559-02

To whom it may concern:

Mistras Group Inc. would like to remove the storage and use location of 2549 Enterprise Avenue, Billings, Montana listed in Condition 10A of Radioactive Materials License 12-16559-02. Along with this request I have attached a copy of the decommission wipe test taken of the storage area.

If you should require any further information I can be reached at 713-899-4338.

Thank you,

Daryl Istre
Corporate Radiation Safety Officer
Mistras Group Inc.

RECEIVED APR 04 2015

SUNTRAC SERVICES, INC.
1818 EAST MAIN STREET
LEAGUE CITY, TEXAS 77573
PHONE (281) 338-2133


AREA WIPE TEST REPORT FORM

Company Name: Mistras Group
Smear Taken Date: April 1, 2015 by Randy Bethea

This report certifies that wipe test smears from Mistras Group located at 2549 Enterprise Ave, Billings, MT 59102 have been assayed at our facility for removable contamination with a Ludlum 2929, S/N 152274. Calibration certificate attached.

Smear Location	Alpha DPM	Beta/Gamma DPM
Taken on floor in front of storage box	BDL	BDL
Taken on the wall behind the storage box	BDL	BDL

CERTIFIED BY:




Jessica Ochoa

DATE:

4.6.15

April 6, 2015

QC'D BY:



Brenda Wasek

DATE:

4-6-15

April 6, 2015

BDL = Below Detectable Limits
DPM = Disintegrations Per Minute



SUNTRAC Services, Inc.
CERTIFICATE OF CALIBRATION

1818 East Main Street
 League City, TX 77573
 Phone: (281) 338-2133
 Fax: (281) 338-2136
 www.suntrac.com

Customer: 147
 Suntrac - IN HOUSE METERS
 1818 East Main Street
 League City, TX 77573

Instrument Serviced:		Manufacturer	Model Number	Serial Number		
Detector:		Ludlum	2929 (L)	152274		
Distance Calculated (inches)		Ludlum	43-10-1	PR156595		
Meter Range Multiplier	Distance Calculated (inches)	Filter	Calculated Calibration Point	Instrument Meter Reading	Unit	Correction Factor
HV	Pulser	N	500	500	Volts	1.0
HV	Pulser	N	2000	2000	Volts	

Calibrated in accordance with Texas Radioactive Materials License Number L03062, and requirements of ANSI/NCSL Z540-1-1994; ANSI-N323A-1997; and 10 CFR 34.25 & 35.61. NIST Traceable #141271/T140856.

Temperature: 79°F Humidity: 41% Altitude: 30inHG Geotropism:
 Calibration Source S/N: A280 Source Activity: 58.07 mCi Cs137: 19.67 mR/hr.@100cm;<+/-5%
 Pulser Utilized: Yes Pulser Information: Ludlum Model 500 S/N 50802

Battery Check: Chargeable OP Volts: 900 Detector Position: Parallel
 Check Source: No Isotope: Isotope S/N: Volts At Contact:

Technician Comments: This instrument is calibrated electronically in C. P. M.

Benito Herrera

4/2/15
 Calibrated By Benito Herrera Cal. Date

6 months 10/2/15
 Calibration Interval Cal. Due Date



SUNTRAC Services, Inc.
CERTIFICATE OF CALIBRATION

1818 East Main Street
 League City, TX 77573
 Phone: (281) 338-2133
 Fax: (281) 338-2136
 www.suntrac.com

Customer: 147
 Suntrac - IN HOUSE METERS
 1818 East Main Street
 League City, TX 77573

Instrument Serviced:		Manufacturer	Model Number	Serial Number		
Detector:		Ludlum	2929 (L)	152274		
Distance Calculated (inches)		Ludlum	43-10-1	PR156595		
Meter Range Multiplier	Distance Calculated (inches)	Filter	Calculated Calibration Point	Instrument Meter Reading	Unit	Correction Factor
Alpha Channel	Pulser	N	40	41	C.P.M.	1.0
Alpha Channel	Pulser	N	400	405	C.P.M.	
Alpha Channel	Pulser	N	4000	4056	C.P.M.	
Alpha Channel	Pulser	N	40000	40566	C.P.M.	
Alpha Channel	Pulser	N	400000	405364	C.P.M.	
Beta/Gamma Cf	Pulser	N	40	41	C.P.M.	1.0
Beta/Gamma Cf	Pulser	N	400	405	C.P.M.	
Beta/Gamma Cf	Pulser	N	4000	4039	C.P.M.	
Beta/Gamma Cf	Pulser	N	40000	40482	C.P.M.	
Beta/Gamma Cf	Pulser	N	400000	404405	C.P.M.	
Minutes	Pulser	N	0.1	0.1	Min	1.0
Minutes	Pulser	N	1	1	Min	
Minutes	Pulser	N	10	10	Min	

Calibrated in accordance with Texas Radioactive Materials License Number L03062, and requirements of ANSI/NCSL Z540-1-1994; ANSI-N323A-1997; and 10 CFR 34.25 & 35.61. NIST Traceable #141271/T140856.

Temperature: 79°F Humidity: 41% Altitude: 30inHG Geotropism:

Calibration Source S/N: A280 Source Activity: 58.07 mCi Cs137: 19.67 mR/hr.@100cm;<+/-5%

Pulser Utilized: Yes Pulser Information: Ludlum Model 500 S/N 50802

Battery Check: Chargeable OP Volts: 900 Detector Position: Parallel

Check Source: No Isotope: Isotope S/N: Min At Contact:

Technician Comments: This instrument is calibrated electronically in C. P. M.

Benito Herrera

4/2/15
 Calibrated By Benito Herrera Cal. Date

6 months 10/2/15
 Calibration Interval Cal. Due Date

D I strc
3810 Happy Hollow
Montgomery, TX
77356



U.S. NRC Region 111
2443 Warnerville Ro. Ste 210
Lisle, IL 60532-4352

605324352

