

## **Parker, Bryan**

---

**From:** Tracy King <tking@mpcphysics.com>  
**Sent:** Tuesday, April 21, 2015 5:24 PM  
**To:** Parker, Bryan  
**Subject:** RE: Question on Covenant Medical renewal

Bryan,

I think that the Mentor seed may no longer be available. This was a carryover from the previous license. Please just drop the Mentor seed.

I am trying to find my notes about the users and not having any luck. Please just list the other two which you located and I will review the entire list at the next RSC meeting.

Thank you so very much for your help and patience.

tracy

**From:** Parker, Bryan [mailto:[Bryan.Parker@nrc.gov](mailto:Bryan.Parker@nrc.gov)]  
**Sent:** Thursday, April 16, 2015 5:22 PM  
**To:** [tking@mpcphysics.com](mailto:tking@mpcphysics.com)  
**Subject:** Question on Covenant Medical renewal

Hey Tracy,

I have a question on the Covenant Medical Center NRC license renewal. I need you to verify/clarify one of the sealed sources listed for 35.400.

Currently on the license (and in the application dated 11/24/14 – Item 6.D.), there is a sealed source noted as **“Mentor Model I-125 SL or I-125 SH.”**

I don't find this source in the SS&D and I'm hoping you can shed some light on it. The model number is remarkably similar to the source noted in Item 6.F. for 35.1000 – “Core Oncology Inc., (formerly Mills Biopharmaceuticals, LLC) Model No. I-125 SL (ProstaSeed)”; however, there is no mention of “Mentor” in the Core Oncology SS&D sheet.

Can you check on this and see if it is needed or know how it should be designated?

Thanks.  
Bryan

*Bryan A. Parker*  
Health Physicist  
USNRC Region III  
[bryan.parker@nrc.gov](mailto:bryan.parker@nrc.gov)  
678-828-7050  
630-515-1078 (fax)

