

**MICHIGAN STATE  
UNIVERSITY**

April 30, 2015

To: Sara Forster  
Health Physicist Licensing Reviewer  
U.S. Nuclear Regulatory Commission – Region III  
Division of Nuclear Materials Safety  
2443 Warrenville Rd. – Ste. 210  
Lisle, IL 60532

From: Jean Chisnell  
Radiation Safety Officer  
Michigan State University

Re: Mail Control Number 585973  
License 21-00021-29

Thorium 228 Sealed Source 1 mCi total possession limit;  
No single source to exceed 10 microCuries;  
Used for calibration



**Vice President  
for Research  
and Graduate  
Studies**

Office of  
Environmental  
Health & Safety

Engineering Research  
Complex  
1449 Engineering  
Research Court, C-125  
East Lansing, MI 48824

517-355-0153  
Fax: 517-353-4871  
ehs.msu.edu

*Jean A Chisnell  
RSO  
Michigan State University*

## Forster, Sara

---

**From:** Chisnell, Jean <jchisnel@ehs.msu.edu>  
**Sent:** Thursday, April 30, 2015 12:21 PM  
**To:** Forster, Sara  
**Subject:** RE: Mail Control Number 585973; License 21-00021-2  
**Attachments:** Amendment Clarification Letter 4-2015.pdf

Ms. Forster,  
The letter of clarification is attached. If you need any additional information please contact me.

Thank you for your assistance,

*Jean*

Jean Chisnell  
Radiation Safety Officer  
Environmental Health & Safety  
Michigan State University

Engineering Research Complex  
1449 Engineering Research Court, C-125  
East Lansing, MI48824  
Phone: 517-355-6734  
FAX: 517-353-4871  
[jchisnel@msu.edu](mailto:jchisnel@msu.edu)  
<http://www.ehs.msu.edu>

---

**From:** Forster, Sara [<mailto:Sara.Forster@nrc.gov>]  
**Sent:** Thursday, April 30, 2015 10:56 AM  
**To:** Chisnell, Jean  
**Subject:** RE: Mail Control Number 585973; License 21-00021-2

Hi Ms. Chisnell:

Will you be able to submit the clarification letter, regarding the thorium-228 sealed sources (not to exceed 10 microcuries per source, and listing a maximum possession limit)? If not today, please let me know when you will be able to submit the response.

Thank you in advance,

Sara A. Forster, Health Physicist Licensing Reviewer  
U.S. Nuclear Regulatory Commission - Region III  
Division of Nuclear Materials Safety  
2443 Warrenville Rd. - Ste. 210  
Lisle, IL 60532-4352  
[sara.forster@nrc.gov](mailto:sara.forster@nrc.gov)  
Direct: (630) 829-9892