

**From:** [Elliott, Robin](#)  
**To:** [Sharon Long](#)  
**Cc:** [kstoneberg@nationalphysics.com](mailto:kstoneberg@nationalphysics.com)  
**Subject:** Termination of U.S. NRC License No. 47-31020-01  
**Date:** Tuesday, April 07, 2015 4:17:00 PM

---

Docket No. 03036877  
License No. 47-31020-01

SUBJECT: TRI-STATE CARDIOLOGY, REQUEST FOR ADDITIONAL INFORMATION CONCERNING APPLICATION FOR AMENDMENT TO LICENSE, CONTROL NO.

Dear Ms. Long:

This is in reference to the letter dated March 1, 2015 requesting to terminate Nuclear Regulatory Commission License No. 47-31020-01. In order to continue our review, we need the following additional information:

Please submit a complete list of the sealed sources in the inventory that will be transferred from Tri-State Cardiology to Reynolds Memorial Hospital, Inc. at the time of the termination of Tri-State Cardiology. In addition, include a copy of the last leak test results for these sources.

Current NRC regulations and guidance are included on the NRC's website at [www.nrc.gov](http://www.nrc.gov); select **Nuclear Materials; Med, Ind, & Academic Uses**; then **Licensee Toolkits, see our toolkit index page**. You may also obtain these documents by contacting the Government Printing Office (GPO) toll-free at 1-866-512-1800. The GPO is open from 8:00 a.m. to 5:30 p.m. EST, Monday through Friday (except Federal holidays).

We will continue our review upon receipt of this information. Please reply to my attention at the Region I Office and include the license and docket number on your communication. In addition, the response must be signed by Dr. Dharawat. You may send a letter, facsimile, or email, if you are able to attach a scanned letter signed by Dr. Dharawat. If you have any technical questions regarding this deficiency letter, please call me at (610) 337-5076.

If we do not receive a reply from you within 30 calendar days from the date of this letter, we will assume that you do not wish to pursue your application.

Regards,

*Robin L. Elliott*

Health Physicist  
U. S. Nuclear Regulatory Commission  
Region I, Division of Nuclear Materials Safety  
2100 Renaissance Blvd  
King of Prussia, PA 19406-2713  
(610) 337-5076 voice

(610) 337-5269 fax

[Robin.Elliott@nrc.gov](mailto:Robin.Elliott@nrc.gov)