



UNITED STATES  
**NUCLEAR REGULATORY COMMISSION**  
 REGION IV  
 1600 EAST LAMAR BLVD  
 ARLINGTON, TEXAS 76011-4511

EMAIL



**Name:** Steven E. Liston, M.D./John Rosenbaum, R.T. License: 50-17838-01  
 Docket: 030-13426  
 Control: 585345

**Organization:** Northern Montana Hospital  
**Phone:** 406-265-2211/406-262-1221  
**E-mail Address:** sliston@pacbell.com/rosejohc@nmhcare.org  
**From:** Jacqueline D. Cook  
**Date:** June 2, 2015  
**Subject:** Application dated December 1, 2014 for License Renewal  
**Pages:** 2 (19 with Tables C.2. and C.3. attached)

Dr. Liston/Mr. Rosenbaum

Per your application dated December 1, 2014, the items on the next page are a request for additional information (deficiency) which require your response. **Please respond to this e-mail by Friday, June 5, 2015. I apologize for the short turnaround time; however, you did not submit a complete renewal package.** Our fax number is (817) 200-8263. You may respond by e-mail in pdf format if you'd like. My email address is [Jackie.Cook@nrc.gov](mailto:Jackie.Cook@nrc.gov). When responding to this e-mail, please include the license, docket and control numbers located at the top of this page.

Thanking you in advance for your cooperation, assistance, and prompt response in this matter.

/RA/  
 Jacqueline D. Cook  
 Senior Health Physicist

~~PUBLIC  
 Immediate Release  
 Normal Release  
 A.3 Sensitive-Security Related  
 A.7 Sensitive Internal  
 Other:  
 Date: \_\_\_\_\_  
 Reviewer: \_\_\_\_\_  
 Date: 6/2/15~~

1. Please note that NUREG-1556, Vol. 9, *Consolidated Guidance About Materials Licenses: Program-Specific Guidance About Medical Use Licenses*, has been revised twice and the current revision is 2. Please complete and resubmit tables C.2 and C.3 of NUREG-1556, Vol. 9, Rev. 2 (<http://www.nrc.gov/reading-rm/doc-collections/nuregs/staff/sr1556/v9/r2/>) to reflect the revised guidance. In addition, please update your attachments and references to the various appendices submitted in your renewal application, NRC Form 313 dated December 1, 2014, accordingly using NUREG-1556, Vol. 9, Rev. 2 as guidance.
2. Please submit an updated delegation of authority for Dr. Steven E. Liston as Radiation Safety Officer. You can use the model procedure found in Appendix I of NUREG-1556, Vol. 9, Rev. 2.
3. Please confirm that Dr. Liston and Dr. Earl R. Harrison, Jr. should still be listed as authorized users on your license. If you have any additional authorized users who should be added, please submit their training and experience as necessary using NRC Form 313(aud and/or aut) (<http://www.nrc.gov/reading-rm/doc-collections/forms/>) and the applicable training and experience regulation found in 10 CFR 35.190, 35.290, 35.390, 35.392, 35.394, and/or 35.396, as appropriate. (<http://www.nrc.gov/reading-rm/doc-collections/cfr/part035/>).
4. The diagram you submitted has a numbers on it. Please explain the number leger.

In addition, please expand your diagram to include the following:

- A. Scale used;
- B. Location, room numbers and principal use of each room or area where byproduct material is prepared, used or stored; and
- C. Location, room numbers, and principal use of each adjacent room, including areas above, beside, and below therapy treatment rooms; indicate whether the room is a restricted or unrestricted area as defined in 10 CFR 20.1003

**PUBLIC**

- Immediate Release
- Normal Release

**NON-PUBLIC**

- A.3 Sensitive-Security Related
- A.7 Sensitive Internal
- Other:

Reviewer: JAL Date: 8/2/15