



May 18, 2015

Br. 1

U.S. Nuclear regulatory Commission
Materials Licensing Branch
Region I
475 Allendale Road
King of Prussia, PA 19406

03011644

RE: Radioactive Material License No. 47-16720-01
Preston Memorial Hospital

Gentlemen:

We have completed the transition to the new hospital. All radioactive materials have been transferred to the new facility. A close out survey of our old hospital has been completed and is attached to this amendment. Please remove the old facility located at 300 South Price St. Kingwood, WV from our license.

If you have any questions, please do not hesitate to contact Mark Perna at (412) 551-9259 or markperna@mac.com.

Sincerely,

Melissa Lockwood
President and CEO

586998

NMSS/RGN1 MATERIALS-002



May 18, 2015

U.S. Nuclear regulatory Commission
Materials Licensing Branch
Region I
475 Allendale Road
King of Prussia, PA 19406
RE: Radioactive Material License No. 47-16720-01

Preston Memorial Hospital

To Whom It May Concern:

We have decommissioned our old hospital. The following is a close-out survey performed on May 16, 2015 by our Chief Technologist Missy Wilt and our Radiation Safety Officer Mark Perna.

A single gross contamination smear was taken on all walls and on the counters, floors, doors and door knobs. The smear was counted in a Capintec CRC 25W with an efficiency for Tc-99m measured at 92.55% on 5/15/15 and minimum detectable activity of 110 for Co-57/Tc-99m and 120 for Cs-137. Area radiation level surveys were taken at each smear site. The background for the area surveys was 0.02 mR/hr using a Ludlum 14C survey meter last calibrated on April 3, 2015. All readings were 0.02 mR/hr.

The following are the smear results for each area tested. Please refer to the enclosed diagram. Background was 417 cpm.

area # DPM above background
gross smear 101

The results are expected as the only unsealed isotope recently used was Tc-99 and the last use of the isotope was more than a week ago. The sealed sources were routinely leaked tested with a negative result.

We will release the area for unrestricted use upon hearing from your office.

Sincerely,

Melissa Lockwood
President and CEO

This is to acknowledge the receipt of your letter/application dated

5/18/15, and to inform you that the initial processing which includes an administrative review has been performed.

Amendment (47-16720-01)
There were no administrative omissions. Your application was assigned to a technical reviewer. Please note that the technical review may identify additional omissions or require additional information.

Please provide to this office within 30 days of your receipt of this card

A copy of your action has been forwarded to our License Fee & Accounts Receivable Branch, who will contact you separately if there is a fee issue involved.

Your action has been assigned **Mail Control Number** 586998.
When calling to inquire about this action, please refer to this control number.
You may call us on (610) 337-5398, or 337-5260.