

U.S. NUCLEAR REGULATORY COMMISSION**Date:** May 29, 2015**TELEPHONE CONVERSATION RECORD**

Mail Control No.	586034	License No.	47-09772-02	Docket No.	030-03390
-------------------------	---------------	--------------------	--------------------	-------------------	------------------

Name of Licensee: Connecticut MultiSpecialty Group, PC

Name of Participant(s): Ray Carlson, RSO
Shawn Seeley, Health Physicist, RI DNMS

Telephone No. 734-395-7361

Subject: Follow up Questions to Renewal Application
(NOTE: This will be used as the Documents Title in ADAMS)

I called Mr. Carlson to get clarification on two items with their renewal application.

1: I mentioned that the application did not specifically mention what type of probe was utilized with their survey instrument. He mentioned that it was an end window GM type.

2: I also mentioned that the application did not adequately describe the areas immediately above and below the nuclear use areas. He mentioned that the suites above and below were standard offices with typical office type occupancy during the week.

I thanked him for the information and told him he should expect his renewed license sometime in the next couple of weeks.

Document Availability: **Publicly Available** **Non-Publicly Available**
 Non-Sensitive
 Sensitive – Privacy Act/ PII
 Immediate Release
 Non-Sensitive Copyright
 Sensitive – Internal
Normal Release Date:
 Sensitive- Proprietary
 Sensitive – Security-Related
Delayed Release Date:
SUNSI Review Completed By: SSeeley / RA /

Document Accession #: _____