

License: 50-29114-01
Docket: 030-34496
Control: 586616

Cook, Jackie

From: Cook, Jackie
Sent: Thursday, June 18, 2015 11:46 AM
To: 'kenneth.harper@yahoo.com'
Subject: Blood Bank of Alaska, Inc. License No. 50-29114-01 Amendment Request dated March 27, 2015

Importance: High

Mr. Harper: Per the direction of Mr. Robert Scanlon, CEO, of Blood Bank of Alaska, Inc., I was told to direct questions/comments regarding the amendment request to name you as Radiation Safety Officer (RSO) for their self-shielded blood irradiator. In the amendment request dated March 27, 2015, it was stated that you are/has been named RSO on a number of broad scope medical licenses. However, the license submitted with the amendment request was the Colorado Department of Public Health & Environment Radioactive Materials License for Mercy Regional Medical Center which appears to authorize only uptake, dilution, or excretion studies which is not adequate. Please submit a copy of one of the broad scope medical licenses listing you as an RSO for which the same/similar manufacturer and model self-shielded irradiator that Blood Bank of Alaska, Inc., License No. 50-29114-01, is authorized. Please note that NUREG-1556, Vol. 5, *Consolidated Guidance About Materials Licenses: Program-Specific Guidance About Self-Shielded Irradiator Licenses*, dated October 1998 (<http://www.nrc.gov/reading-rm/doc-collections/nuregs/staff/sr1556/v5/>) is the current guidance used for self-shielded irradiator licensees, specifically 8.7 Item 7, 8.7.1.

Please submit a copy of the broad scope medical license with self-shielded irradiator authorization listing you as the RSO or submit an applicable response in "response from applicant" section of NUREG-1556, Vol. 5, Item 7, 8.7.1 as soon as practical but no later than Monday, June 22, 2015. In addition, please include Mr. Scanlon in your email response so he can be aware of the request for additional information.

Please do not hesitate to contact me at your convenience if I can be of further assistance regarding the above.

Sincerely,

Ms. Jacqueline "Jackie" D. Cook
Senior Health Physicist
US Nuclear Regulatory Commission Region IV
Division of Nuclear Materials Safety
Nuclear Materials Safety Branch B
1600 East Lamar Blvd., Arlington, TX 76011
817-200-1132 (office)/817-200-1263 (fax)
Email address: Jackie.Cook@nrc.gov

PUBLIC

- Immediate Release
- Normal Release

NON-PUBLIC

- A.3 Sensitive-Security Related
- A.7 Sensitive Internal
- Other:

Reviewer: JPC Date: 6/10/15