

From: [Lawyer, Dennis](#)
To: Gary.C.Brown-2@chemours.com
Subject: E.I. DuPont De Nemours and Company, Inc., Request for Additional Information Concerning Application for a License Renewal, Control 586298
Date: Wednesday, June 17, 2015 3:43:00 PM

Dear Mr. Gary C. Brown,

This is in reference to your application dated February 27, 2015, and additional letter dated June 11, 2015, requesting for renewal to Nuclear Regulatory Commission License No. 07-00455-29, Docket No. 03003861. In order to continue our review, we need the following additional information:

1. Please confirm that the LIXI device is a Model LS-82X with Source MDS Nordion Model C-381. This unit appears to be a portable gauge which requires the additional information:
 - a. Please provide the statement: "Before using licensed materials, authorized users will have successfully completed one of the training courses described in Criteria in the section entitled 'Training for Individuals Working In or Frequenting Restricted Areas' in NUREG-1556, Vol. 1, Rev. 1, 'Consolidated Guidance about Materials Licenses: Program-Specific Guidance about Portable Gauge Licenses,' dated November 2001"; or A description of the training and experience for proposed gauge users.
 - b. Please provide a statement: "We will implement and maintain the operating and emergency procedures in ERRATA (<http://pbadupws.nrc.gov/docs/ML0521/ML052130055.pdf>) to Appendix H dated July 5, 2005, of NUREG-1556, Vol. 1, Rev. 1, 'Consolidated Guidance about Materials Licenses: Program-Specific Guidance about Portable Gauge Licenses,' dated November 2001, and provide copies of these procedures to all gauge users and at each job site," or submit alternative procedures.
2. Previously in License Condition 18, you were authorized to perform some non-routine maintenance. You did not request for this authorization in your current applications. Please confirm that you do not wish to perform non-routine maintenance or provide the information listed in Appendix N of NUREG-1556, Volume 4, supporting a request to perform non-routine operations in-house.
3. Please note that currently your license is planned to be transferred to the Chemours Company, F.C., L.L.C. Please have a management representative from the Chemours Company sign that they have read and concurred with your application and please have them sign concurring with all additional information submitted associated with this license. If the Plant Manager, Elizabeth Schowe, is authorized to sign for both the licensee and Chemours Company, F.C., L.L.C., please make that statement.

We will continue our review upon receipt of this information. Please reply to my attention at the Region 1 Office (Address below) and refer to Mail Control No. <No>. If you have

technical questions regarding this letter, please call me at (610) 337-5366.

Please note that you may not reply to this letter by return e-mail. Your reply must be in writing by letter, facsimile (610-337-5269), or signed letter attached to an email. If we do not receive a reply from you within 30 calendar days from the date of this e-mail, we will assume that you do not wish to pursue your application.

Region 1 Office Mailing Address: Licensing Assistance Team, US Nuclear Regulatory Commission Region I, 2100 Renaissance Boulevard, Suite 100, King of Prussia, PA 19406-2713.

Dennis Lawyer
Health Physicist
U.S. Nuclear Regulatory Commission
Division of Nuclear Material Safety
610-337-5366
610-337-5269 (F)