

Docket File Information
SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: Cargill, Inc. 1503 Wabash Avenue Lafayette, IN 47905 REPORT NUMBER(S) 2015-001	2. NRC/REGIONAL OFFICE Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352
--	---

3. DOCKET NUMBER(S) 030-28833	4. LICENSE NUMBER(S) 13-24550-01	5. DATE(S) OF INSPECTION July 23, 2015
--------------------------------------	---	---

6. INSPECTION PROCEDURES USED 87124	7. INSPECTION FOCUS AREAS 03.01-03.07
--	--

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S) 03120	2. PRIORITY 5	3. LICENSEE CONTACT Allen Crose, RSO	4. TELEPHONE NUMBER (765) 420-6610
---------------------------------	----------------------	---	---

Main Office Inspection Next Inspection Date: 7/23/2020

Field Office Inspection _____

Temporary Job Site Inspection _____

PROGRAM SCOPE

This was a routine inspection of the Cargill, Inc., facility in Lafayette, Indiana. This facility employed 40 individuals, operated continuously, and processed soy beans for feed products. The licensee possessed 6 gauging devices containing cesium-137 sources throughout its plant and utilized these gauges for process measurements. Service activities and leak testing were performed by the device manufacturer.

Performance Observations

During this inspection, the inspector toured the licensee's plant and noted the environmental conditions, reviewed select records, and interviewed select licensee personnel. Licensed material was adequately secured and not readily accessible to members of the general public.

The inspector performed independent radiation surveys around the devices. These measurements did not indicate readings in excess of Title 10 of the Code of Federal Regulations (10 CFR) Part 20 limits in restricted or unrestricted areas. The inspector noted all sealed source leak tests, inventories and shutter tests were performed at the required frequencies.

The inspector noted that during the period since the previous inspection, the licensee transferred one fixed gauging device to the manufacturer. The inspector verified that the licensee maintained appropriate transfer records documenting this transfer as well as records verifying receipt by the manufacturer.

No violations were identified during this inspection.