

1053 Sehome Ave. Bellingham, WA 98229-8866  
Ph.: 800-252-7896 Fax: 360-671-8686  
E-mail: [cjbruyn@outlook.com](mailto:cjbruyn@outlook.com)  
[www.cjbruynco.com](http://www.cjbruynco.com)

---

U.S. Nuclear Regulatory Commission  
Attn: GLTS  
11555 Rockville Pike  
Rockville, MD 20852

**RE: Quarterly Reports of Transfer of Ownership – 2015 / Quarter 2  
WA. State Specific License: WN-I0523-1**

7/31/2015


Dear Sir,

Enclosed is documentation collated by Region for ownership transfers of Electron Capture Detectors containing the isotope Ni63 issued for the quarter and region stated above.

Any questions or comments can be addressed to Caline J. Bruyn, RSO - CJ Bruyn & Co. at the address or phone and fax numbers indicated above.

Salutations,

Caline J. Bruyn



1053 Sehome Ave. Bellingham, WA 98229-8866  
Ph.: 800-252-7896 Fax: 360-671-8686  
E-mail: cjbruyn@msn.com

[www.cjbruynco.com](http://www.cjbruynco.com)

---

U.S. Nuclear Regulatory Commission  
Attn: GLTS  
11555 Rockville Pike  
Rockville, MD 20852

RE: Quarterly Reports of Transfer of Ownership – 2015 / Qtr 4. Region 1  
WA. State Specific License: WN-I0523-1

7/31/2015


Dear Sir,

Enclosed is documentation for ownership transfers of Electron Capture Detectors containing the isotope Ni63 issued for the quarter and region stated above.

Any questions or comments can be addressed to Caline J. Bruyn, RSO - CJ Bruyn & Co. at the address or phone and fax numbers indicated above.

Salutations,

  
Caline J. Bruyn



---

1053 Sehome Ave., Bellingham, Wa 98229-8866 U.S.A.  
800-252-7896 Ph.: 360-671-0182 Fax: 360-671-8686 E-mail: cjbruyn@msn.com Website: www.cjbruynco.com

---

PA Bureau of Radiation  
Dept. of Environmental Protection  
P.O. Box 8469  
Harrisburg, PA 97232-2162

5/15/2015

~~Out of state sale, exempt from sales tax~~

Dear Sir,

**This letter is to inform you that CJ Bruyn & Co., with authorization of the Washington State Radiologic Health Section, under license # WN-I0523-1, has assisted in transferring ownership of the following Generally Licensed Device(s):** Electron Capture Detector (ECD) for use in a Gas Chromatograph (GC)

Date of Transfer: 5/15/2015                      2 Each                      S/N(s): L1956 F4916

Manufactured by: Agilent                      Model: 19233                      Isotope: <15 Milli Curie Ni63


Prior Owner:                      Morgan Schaffer  
8300 St. Patrick, Suite 150  
LaSalle, Quebec CANADA H8n 2H1  
Attn: Brenton Grant

New Owner                      Morgan Schaffer USA  
101 Henderson Dr.  
Sharon Hill, PA 19079  
Bill Zuccarello

The ECD has been surveyed at or near the date of transfer. The new user / owner has been provided with a Leak-Test Record. We have verified that the New User / Owner is in possession of a Radiation Safety manual and / or RSO Information Letter and has been reminded that the operator is required to read it and stay familiar with the regulations. We have provided the New owner with the NRC regulations regarding ownership of a generally licensed device. CJ Bruyn & Co. has made sure that the labels are legible and firmly attached to the ECDs.

Signed for authenticity by,

CJ Bruyn, RSO



1053 Sehome Ave. Bellingham, WA 98229-8866  
Ph.: 800-252-7896 Fax: 360-671-8686  
E-mail: [cjbruyn@msn.com](mailto:cjbruyn@msn.com)

---

U.S. Nuclear Regulatory Commission  
Attn: GLTS  
11555 Rockville Pike  
Rockville, MD 20852

RE: Quarterly Reports of Transfer of Ownership – 2015 / Qtr. 2 /Region 2  
WA. State Specific License: WN-I0523-1

7/31/2015


Dear Sir,

There were no ownership transfers of Electron Capture Detectors containing the isotope Ni63 issued for the quarter and region stated above.

Any questions or comments can be addressed to Caline J. Bruyn, RSO - CJ Bruyn & Co. at [cjbruyn@msn.com](mailto:cjbruyn@msn.com) or the address, phone and fax numbers indicated above.

Salutations,

Caline J. Bruyn



1053 Sehome Ave. Bellingham, WA 98229-8866  
Ph.: 800-252-7896 Fax: 360-671-8686  
E-mail: [cjbruyn@msn.com](mailto:cjbruyn@msn.com)

[www.cjbruynco.co](http://www.cjbruynco.co)

---

U.S. Nuclear Regulatory Commission  
Attn: GLTS  
11555 Rockville Pike  
Rockville, MD 20852

**RE: Quarterly Reports of Transfer of Ownership – 2015 / Qtr. 2 /Region 3  
WA. State Specific License: WN-I0523-1**

7/31/2015

Dear Sir,

There were no ownership transfers of Electron Capture Detectors containing the isotope Ni63 issued for the quarter and region stated above.

Any questions or comments can be addressed to Caline J. Bruyn, RSO - CJ Bruyn & Co. at [cjbruyn@msn.com](mailto:cjbruyn@msn.com) or the address, phone and fax numbers indicated above.

Salutations,



Caline J. Bruyn



1053 Sehome Ave. Bellingham, WA 98229-8866  
Ph.: 800-252-7896 Fax: 360-671-8686  
E-mail: cjbruyn@msn.com

---

U.S. Nuclear Regulatory Commission  
Attn: GLTS  
11555 Rockville Pike  
Rockville, MD 20852

RE: Quarterly Reports of Transfer of Ownership – 2015 / Qtr 4. Region 4  
WA. State Specific License: WN-I0523-1

7/31/2015

Dear Sir,


Enclosed is documentation for ownership transfers of Electron Capture Detectors containing the isotope Ni63 issued for the quarter and region stated above.

Any questions or comments can be addressed to Caline J. Bruyn, RSO - CJ Bruyn & Co. at the address or phone and fax numbers indicated above.

Salutations,

A handwritten signature in black ink, appearing to read 'CJ Bruyn', is written over a printed name.

Caline J. Bruyn



---

1053 Sehome Ave., Bellingham, Wa 98229-8866 U.S.A.  
800-252-7896 Ph.: 360-671-0182 Fax: 360-671-8686 E-mail: cjbruyn@msn.com Website: www.cjbruynco.com

---

Director  
WA DOH Office of Rad Protection  
PO Box 47827  
Olympia, WA 98504-782

5/25/2015

Dear Sir,

**This letter is to inform you that CJ Bruyn & Co., with authorization of the Washington State Radiologic Health Section, under license # WN-I0523-1, has assisted in transferring ownership of the following Generally Licensed Device(s):** Electron Capture Detector (ECD) for use in a Gas Chromatograph (GC)

Date of Transfer: 5/25/2015                      1 Each                      S/N(s): F5253

Manufactured by: Agilent Tech.                      Model: G1224A                      Isotope: <15 Milli Curie Ni63

Prior Owner:                      Test America Portland  
9405 SW Nimbus Ave.  
Beaverton, OR 97008  
Attn: Sandra Langway  
503-906-9200


New Owner                      Test America, Inc. Spokane  
11922 E 1st. Ave.  
Spokane, WA  
Attn: Chris Williams  
509-924-9200

The ECD has been surveyed at or near the date of transfer. The new user / owner has been provided with a Leak-Test Record. We have verified that the New User / Owner is in possession of a Radiation Safety manual and / or RSO Information Letter and has been reminded that the operator is required to read it and stay familiar with the regulations. We have provided the New owner with the NRC regulations regarding ownership of a generally licensed device. CJ Bruyn & Co. has made sure that the labels are legible and firmly attached to the ECDs.

Signed for authenticity by,

A handwritten signature in cursive script that reads 'CJ Bruyn'.

CJ Bruyn, RSO



---

1053 Sehome Ave., Bellingham, Wa 98229-8866 U.S.A.

800-252-7896 Ph.: 360-671-0182 Fax: 360-671-8686 E-mail: [cjbruyn@msn.com](mailto:cjbruyn@msn.com) Website: [www.cjbruynco.com](http://www.cjbruynco.com)

---

OR Health Services  
Dept. of Human Services  
800 NE Oregon St. Suite 640  
Portland, OR 97232-2162  
Attn: Program Director

5/25/2015

Dear Sir,

**This letter is to inform you that CJ Bruyn & Co., with authorization of the Washington State Radiologic Health Section, under license # WN-10523-1, has assisted in transferring ownership of the following Generally Licensed Device(s):** Electron Capture Detector (ECD) for use in a Gas Chromatograph (GC)

Date of Transfer: 5/25/2015                      1 Each                      S/N(s): F5253

Manufactured by: Agilent Tech.                      Model: G1224A                      Isotope: <15 Milli Curie Ni63

Prior Owner:                      Test America Portland  
9405 SW Nimbus Ave.  
Beaverton, OR 97008  
Attn: Sandra Langway  
503-906-9200

New Owner                      Test America, Inc. Spokane  
11922 E 1st. Ave.  
Spokane, WA  
Attn: Chris Williams  
509-924-9200


The ECD has been surveyed at or near the date of transfer. The new user / owner has been provided with a Leak-Test Record. We have verified that the New User / Owner is in possession of a Radiation Safety manual and / or RSO Information Letter and has been reminded that the operator is required to read it and stay familiar with the regulations. We have provided the New owner with the NRC regulations regarding ownership of a generally licensed device. CJ Bruyn & Co. has made sure that the labels are legible and firmly attached to the ECDs.

Signed for authenticity by,

A handwritten signature in cursive script that reads 'CJ Bruyn'.

CJ Bruyn, RSO





1053 Sehome Ave., Bellingham, Wa 98229-8866 U.S.A.

800-252-7896 Ph.: 360-671-0182 Fax: 360-671-8686 E-mail: [cjbruyn@msn.com](mailto:cjbruyn@msn.com) Website: [www.cjbruynco.com](http://www.cjbruynco.com)

Gary Robertson, Director  
WA DOH Office of Rad Protection  
PO Box 47827  
Olympia, WA 98504-782

5/25/2015

Dear Sir,

**This letter is to inform you that CJ Bruyn & Co., with authorization of the Washington State Radiologic Health Section, under license # WN-I0523-1, has assisted in transferring ownership of the following Generally Licensed Device(s):** Electron Capture Detector (ECD) for use in a Gas Chromatograph (GC)

Date of Transfer: 5/25/2015                      1 Each                      S/N(s): L6647

Manufactured by: Agilent Tech.                      Model: 19233                      Isotope: <15 Milli Curie Ni63


Prior Owner:                      Test America Portland  
9405 SW Nimbus Ave.  
Beaverton, OR 97008  
Attn: Sandra Langway  
503-906-9200

New Owner                      Test America, Inc. Spokane  
11922 E 1st. Ave.  
Spokane, WA  
Attn: Chris Williams  
509-924-9200

The ECD has been surveyed at or near the date of transfer. The new user / owner has been provided with a Leak-Test Record. We have verified that the New User / Owner is in possession of a Radiation Safety manual and / or RSO Information Letter and has been reminded that the operator is required to read it and stay familiar with the regulations. We have provided the New owner with the NRC regulations regarding ownership of a generally licensed device. CJ Bruyn & Co. has made sure that the labels are legible and firmly attached to the ECDs.

Signed for authenticity by,

CJ Bruyn, RSO



---

---

1053 Sehome Ave., Bellingham, Wa 98229-8866 U.S.A.

800-252-7896 Ph.: 360-671-0182 Fax: 360-671-8686 E-mail: [cjbruyn@msn.com](mailto:cjbruyn@msn.com) Website: [www.cjbruynco.com](http://www.cjbruynco.com)

---

---

OR Health Services  
Dept. of Human Services  
800 NE Oregon St. Suite 640  
Portland, OR 97232-2162  
Attn: Program Director

5/25/2015

Dear Sir,

**This letter is to inform you that CJ Bruyn & Co., with authorization of the Washington State Radiologic Health Section, under license # WN-I0523-1, has assisted in transferring ownership of the following Generally Licensed Device(s):** Electron Capture Detector (ECD) for use in a Gas Chromatograph (GC)

Date of Transfer: 5/25/2015                      1 Each                      S/N(s): L6647

Manufactured by: Agilent Tech.                      Model: 19233                      Isotope: <15 Milli Curie Ni63

Prior Owner:                      Test America Portland  
9405 SW Nimbus Ave.  
Beaverton, OR 97008  
Attn: Sandra Langway  
503-906-9200


New Owner                      Test America, Inc. Spokane  
11922 E 1st. Ave.  
Spokane, WA  
Attn: Chris Williams  
509-924-9200

The ECD has been surveyed at or near the date of transfer. The new user / owner has been provided with a Leak-Test Record. We have verified that the New User / Owner is in possession of a Radiation Safety manual and / or RSO Information Letter and has been reminded that the operator is required to read it and stay familiar with the regulations. We have provided the New owner with the NRC regulations regarding ownership of a generally licensed device. CJ Bruyn & Co. has made sure that the labels are legible and firmly attached to the ECDs.

Signed for authenticity by,

A handwritten signature in cursive script that reads 'CJ Bruyn'.

CJ Bruyn, RSO



---

1053 Sehome Ave., Bellingham, Wa 98229-8866 U.S.A.  
800-252-7896 Ph.: 360-671-0182 Fax: 360-671-8686 E-mail: cjbruyn@msn.com Website: www.cjbruynco.com

---

OR Health Services  
Dept. of Human Services  
800 NE Oregon St. Suite 640  
Portland, OR 97232-2162  
Attn: Program Director

5/25/2015

Dear Sir,

**This letter is to inform you that CJ Bruyn & Co., with authorization of the Washington State Radiologic Health Section, under license # WN-I0523-1, has assisted in transferring ownership of the following Generally Licensed Device(s):** Electron Capture Detector (ECD) for use in a Gas Chromatograph (GC)

Date of Transfer: 5/25/2015                      2 Each                      S/N(s): U8544 U0788

Manufactured by: Agilent Tech.                      Model: G2397A                      Isotope: <15 Milli Curie Ni63

Prior Owner:                      Test America Portland  
9405 SW Nimbus Ave.  
Beaverton, OR 97008  
Attn: Sandra Langway  
503-906-9200


New Owner                      Test America, Inc. Spokane  
11922 E 1st. Ave.  
Spokane, WA  
Attn: Chris Williams  
509-924-9200

The ECD has been surveyed at or near the date of transfer. The new user / owner has been provided with a Leak-Test Record. We have verified that the New User / Owner is in possession of a Radiation Safety manual and / or RSO Information Letter and has been reminded that the operator is required to read it and stay familiar with the regulations. We have provided the New owner with the NRC regulations regarding ownership of a generally licensed device. CJ Bruyn & Co. has made sure that the labels are legible and firmly attached to the ECDs.

Signed for authenticity by,

A handwritten signature in cursive script that reads 'CJ Bruyn'.

CJ Bruyn, RSO



1053 Sehome Ave., Bellingham, Wa 98229-8866 U.S.A.

800-252-7896 Ph.: 360-671-0182 Fax: 360-671-8686 E-mail: cjbruyn@msn.com Website: www.cjbruynco.com

Director  
WA DOH Office of Rad Protection  
PO Box 47827  
Olympia, WA 98504-782

5/25/2015

Dear Sir,

**This letter is to inform you that CJ Bruyn & Co., with authorization of the Washington State Radiologic Health Section, under license # WN-I0523-1, has assisted in transferring ownership of the following Generally Licensed Device(s):** Electron Capture Detector (ECD) for use in a Gas Chromatograph (GC)

Date of Transfer: 5/25/2015                      2 Each                      S/N(s): U8544 U0788

Manufactured by: Agilent Tech.                      Model: G2397A                      Isotope: <15 Milli Curie Ni63

Prior Owner:                      Test America Portland  
9405 SW Nimbus Ave.  
Beaverton, OR 97008  
Attn: Sandra Langway  
503-906-9200

New Owner                      Test America, Inc. Spokane  
11922 E 1st. Ave.  
Spokane, WA  
Attn: Chris Williams  
509-924-9200

The ECD has been surveyed at or near the date of transfer. The new user / owner has been provided with a Leak-Test Record. We have verified that the New User / Owner is in possession of a Radiation Safety manual and / or RSO Information Letter and has been reminded that the operator is required to read it and stay familiar with the regulations. We have provided the New owner with the NRC regulations regarding ownership of a generally licensed device. CJ Bruyn & Co. has made sure that the labels are legible and firmly attached to the ECDs.

Signed for authenticity by,

CJ Bruyn, RSO



**CJ BRUYN & CO.**  
**ECD Specialist**  
Since 1991

1053 Sehome Ave. Bellingham, WA 98229-8866  
Ph.: 800-252-7896 Fax: 360-671-8686  
E-mail: [cjbruyn@outlook.com](mailto:cjbruyn@outlook.com)

[www.cjbruynco.com](http://www.cjbruynco.com)

---

**Canadian Nuclear Safety Commission**  
280 Slater St.  
PO Box 1046, Station B  
Ottawa, ON K1P 5S9 CANADA

**RE: Quarterly Reports of Transfer of Ownership – 2015 / Quarter 2**  
**WA. State Specific License: WN-I0523-1**

7/31/2015


Dear Sir,

Enclosed is documentation for ownership transfers of Electron Capture Detectors containing the isotope Ni63 issued for the quarter and region stated above.

Any questions or comments can be addressed to Caline J. Bruyn, RSO - CJ Bruyn & Co. at the address or phone and fax numbers indicated above.

Salutations,

Caline J. Bruyn



---

1053 Sehome Ave., Bellingham, Wa 98229-8866 U.S.A.  
800-252-7896 Ph.: 360-671-0182 Fax: 360-671-8686 E-mail: cjbruyn@msn.com Website: www.cjbruynco.com

---

Canada Nuclear Safety Commission  
280 Slater St.  
PO Box 1046, Station B  
Ottawa, ON K1P 5S9 CANADA  
~~Out of state sale, exempt from sales tax~~

5/15/2015

Dear Sir,

**This letter is to inform you that CJ Bruyn & Co., with authorization of the Washington State Radiologic Health Section, under license # WN-I0523-1, has assisted in transferring ownership of the following Generally Licensed Device(s):** Electron Capture Detector (ECD) for use in a Gas Chromatograph (GC)

Date of Transfer: 5/15/2015                      2 Each                      S/N(s): L1956 F4916

Manufactured by: Agilent                      Model: 19233                      Isotope: <15 Milli Curie Ni63

Prior Owner:                      Morgan Schaffer  
8300 St. Patrick, Suite 150  
LaSalle, Quebec CANADA H8n 2H1  
Attn: Brenton Grant

New Owner                      Morgan Schaffer USA  
101 Henderson Dr.  
Sharon Hill, PA 19079  
Bill Zuccarello

The ECD has been surveyed at or near the date of transfer. The new user / owner has been provided with a Leak-Test Record. We have verified that the New User / Owner is in possession of a Radiation Safety manual and / or RSO Information Letter and has been reminded that the operator is required to read it and stay familiar with the regulations. We have provided the New owner with the NRC regulations regarding ownership of a generally licensed device. CJ Bruyn & Co. has made sure that the labels are legible and firmly attached to the ECDs.

Signed for authenticity by,

A handwritten signature in cursive script that reads 'CJ Bruyn'.

CJ Bruyn, RSO