



**UNITED STATES
NUCLEAR REGULATORY COMMISSION**
REGION IV
1600 E. LAMAR BLVD.
ARLINGTON, TX 76011-4511

August 31, 2015

Alan L. Mitchell, M.D., Radiation Safety Officer
Campbell County Memorial Hospital
501 South Burma Avenue
Gillette, Wyoming 82716

SUBJECT: REQUEST FOR ADDITIONAL INFORMATION


The Nuclear Regulatory Commission (NRC) has completed the technical review of the renewal application dated March 17, 2015, response e-mail dated July 27, 2015, letter dated August 5, 2015, and letter dated August 17, 2015 for NRC license number 49-18030-01, and additional information is needed. Please provide your response in a signed and dated letter, in hospital letterhead, within 5 days from receipt of this letter, and make reference to mail control number 586331.

1. Provide the following commitment: "Either we will perform a prospective evaluation demonstrating that unmonitored individuals are not likely to receive, in 1 year, a radiation dose in excess of 10% of the allowable limits in 10 CFR Part 20 or we will provide dosimetry that meets the requirements listed under 'Criteria' in NUREG-1556, Vol. 9, 'Consolidated Guidance about Materials Licenses: Program-Specific Guidance About Medical Use Licenses.'"

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter will be available electronically for public inspection in the NRC Public Document Room or from the NRC's document system (ADAMS). ADAMS is accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>.

Thank you for your cooperation.

Sincerely,



Roberto J. Torres, M.S., Senior Health Physicist
Nuclear Materials Safety Branch B

Docket: 030-14365
License: 49-18030-01
Control: 586331