



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION IV
1600 EAST LAMAR BLVD
ARLINGTON, TEXAS 76011-4511

EMAIL



Name: Ronald Frick License: 53-18126-01
Docket: 030-14529
Organization: Straub Clinic & Hospital Control: 586530
Phone: 808-373-7009(office)
E-mail Address: rfrick@gammacorp.com
From: Jacqueline D. Cook
Date: October 8, 2015
Subject: Application dated March 31, 2015 for License Renewal
Pages: 2

Mr. Frick:

Per your application dated April 6, 2015, the items on the next page are deficiencies which require your response. **I apologize for the short turnaround but if you can please respond to this e-mail by tomorrow, Friday, October 9, 2015, it will be greatly appreciated. If you are unable to respond by tomorrow, please give me a call so we can discuss an alternate date to respond.** Our fax number is (817) 200-1263. You may respond by e-mail in pdf format if you'd like. My email address is Jackie.Cook@nrc.gov. When responding to this e-mail, please include the license, docket and control numbers located at the top of this page.

Please note that your renewal application was reviewed using NUREG-1556, Vol. 9, Rev. 2, "Consolidated Guidance About Materials Licenses: Program-Specific Guidance About Medical Use Licenses" dated January 2008 (<http://www.nrc.gov/reading-rm/doc-collections/nuregs/staff/sr1556/v9/r2/>).

Thanking you in advance for your cooperation, assistance, and prompt response in this matter.

/RA/
Jacqueline D. Cook
Senior Health Physicist

PUBLIC

- Immediate Re:
 Normal Release

NON-PUBLIC

- A.3 Sensitive-Security Related
 A.7 Sensitive Internal
 Other: _____

Reviewer: JDC

Date: 10/9/15

1. Please note that Appendix I of NUREG-1556, Vol.9, Rev. 2 contains a model RSO Delegation of Authority. (<http://www.nrc.gov/reading-rm/doc-collections/nuregs/staff/sr1556/v9/r2/>) Please submit an updated RSO Delegation of Authority for Ronald Frick, M.S., CHP, DABR using Appendix I as guidance.
2. A. In ATT 9.1 A of your renewal application dated April 6, 2015, you state, in part, that all byproduct material stored in the Nuclear Medicine Department will be secured in the Hot Lab or long term storage areas.

We were unable to identify the long term storage area in your drawing.

Please clarify this discrepancy.

- B. Please resubmit your diagrams identifying the following: location, room numbers, and/or principal use of each adjacent room (e.g., office, file, toilet, closet, hallway) including areas above, beside, and below therapy treatment rooms, indicating whether the room is a restricted or unrestricted area as defined in 10 CFR 20.1003.
3. Since you are authorized for 10 CFR 35.300 material, please submit one of the following:
 - A statement that: "Dosages will be determined by relying on the provider's dose label for measurement of the radioactivity and a combination of volumetric measurement and mathematical calculation."

OR

- We are providing a description of the dosage measurement equipment, the nationally recognized calibration standard (or manufacturer's calibration instructions), and dosage measurement procedures.
4. If you are using PET radionuclides, please provide a description of the additional facilities and equipment for these uses.

PUBLIC

- Immediate Release
- Normal Release

NON-PUBLIC

- A.3 Sensitive-Security Related
- A.7 Sensitive Internal
- Other: _____

Reviewer: Date: 10/01/15