



**UNITED STATES
NUCLEAR REGULATORY COMMISSION**
REGION IV
1600 E. LAMAR BLVD
ARLINGTON TX 76011-4511

October 21, 2015

Mr. Tom Carter, Radiology Manager
Marias Medical Center
640 Park Avenue
Shelby, Montana 59474

SUBJECT: TERMINATION OF NRC RADIOACTIVE MATERIALS LICENSE NO. 25-27641-01

By correspondence received August 14, 2015, Kari Cann, Radiation Safety Officer, informed the Nuclear Regulatory Commission (NRC) that Marias Medical Center wished to terminate NRC radioactive materials license number 25-27641-01. The NRC Form 314 dated August 10, 2015, stated that all radioactive material had been transferred to a licensed entity.

The supporting documentation to the NRC Form 314 provided copies of the most recent leak test results demonstrating that there was no evidence of leakage of radioactive material from the sealed sources. In addition, a review of the historical file supported that only radioactive material with decay half-lives of less than 120 days were possessed by the licensee. Based on the documentation submitted and a review of your historical file, the staff concluded that all licensable radioactive material has been removed from 640 Park Avenue, Shelby, Montana, and this location is suitable for unrestricted use in accordance with 10 CFR 20.1402, "Radiological Criteria for Unrestricted Use." Based on these conclusions, no further remediation or actions with respect to NRC regulated material is required.

Please find enclosed Amendment No. 06 to License No. 25-27641-01 authorizing your license termination. An environmental assessment for this action is not required since this action is categorically excluded under 10 CFR 51.22(c)(20)(i) and (ii).

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter and its enclosure will be available electronically for public inspection in the NRC Public Document Room or from the NRC's document system (ADAMS). ADAMS is accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>.

You can contact me at 817-200-1189 if you have questions or require clarification on any of the information stated above.

Sincerely,

/RA/

Roberto J. Torres, M.S., Senior Health Physicist
Nuclear Materials Safety Branch B

Docket: 030-35221
License: 25-27641-01
Control: 588784

Enclosure: As stated