

# GROUP A

FOIA/PA NO: 2016-0068

## RECORDS BEING RELEASED IN PART

The following types of information are being withheld:

- Ex. 1:  Records properly classified pursuant to Executive Order 13526
- Ex. 2:  Records regarding personnel rules and/or human capital administration
- Ex. 3:  Information about the design, manufacture, or utilization of nuclear weapons  
 Information about the protection or security of reactors and nuclear materials  
 Contractor proposals not incorporated into a final contract with the NRC  
 Other \_\_\_\_\_
- Ex. 4:  Proprietary information provided by a submitter to the NRC  
 Other \_\_\_\_\_
- Ex. 5:  Draft documents or other pre-decisional deliberative documents (D.P. Privilege)  
 Records prepared by counsel in anticipation of litigation (A.W.P. Privilege)  
 Privileged communications between counsel and a client (A.C. Privilege)  
 Other \_\_\_\_\_
- Ex. 6:  Agency employee PII, including SSN, contact information, birthdates, etc.  
 Third party PII, including names, phone numbers, or other personal information
- Ex. 7(A):  Copies of ongoing investigation case files, exhibits, notes, ROI's, etc.  
 Records that reference or are related to a separate ongoing investigation(s)
- Ex. 7(C):  Special Agent or other law enforcement PII  
 PII of third parties referenced in records compiled for law enforcement purposes
- Ex. 7(D):  Witnesses' and Allegers' PII in law enforcement records  
 Confidential Informant or law enforcement information provided by other entity
- Ex. 7(E):  Law Enforcement Technique/Procedure used for criminal investigations  
 Technique or procedure used for security or prevention of criminal activity
- Ex. 7(F):  Information that could aid a terrorist or compromise security

Other/Comments: \_\_\_\_\_

ML003718463



UNITED STATES  
NUCLEAR REGULATORY COMMISSION  
REGION I  
475 ALLENDALE ROAD  
KING OF PRUSSIA, PA 19406-1415

May 25, 2000

Docket No. 030-32352  
Control No. 127839

License No. 29-02023-06

Sam I. Brown, M.D.  
Radiation Safety Officer  
Evergreen Radiology Associates, P.A.  
785 Totowa Road  
Totowa, NJ 07512

Dear Dr. Brown:

This refers to your license amendment request. Enclosed with this letter is the amended license.

Please review the enclosed document carefully and be sure that you understand and fully implement all the conditions incorporated into the amended license. If there are any errors or questions, please notify the U.S. Nuclear Regulatory Commission, Region I Office, Licensing Assistance Team, (610) 337-5093 or 5239, so that we can provide appropriate corrections and answers.

Thank you for your cooperation.

Sincerely,

**Original signed by Michelle Beardsley**

Michelle Beardsley  
Health Physicist  
Nuclear Materials Safety Branch 1  
Division of Nuclear Materials Safety

Enclosure:  
Amendment No. 11

A

S. Brown  
Evergreen Radiology Associates, P.A.

DOCUMENT NAME: C:\29-02023-06.127839.05252000.wpd

To receive a copy of this document, indicate in the box: "C" = Copy w/o attach/encl "E" = Copy w/ attach/encl "N" = No copy

OFFICE	DNMS/RI	N	DNMS/RI		DNMS/RI			
NAME	MBeardsley/mrb /s/							
DATE	05/25/2000							

OFFICIAL RECORD COPY

**MATERIALS LICENSE**

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 36, 39, 40, and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below, to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations, and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

<p style="text-align: center;">Licensee</p> <p>1. Evergreen Radiology Associates, P.A.</p> <p>2. 785 Totowa Road Totowa, New Jersey 07512</p>	<p>In accordance with the letter dated February 22, 2000,</p> <p>3. License number 29-02023-06 is amended in its entirety to read as follows:</p> <hr/> <p>4. Expiration date August 31, 2001</p> <hr/> <p>5. Docket No. 030-32352 Reference No.</p>
---	--

6. Byproduct, source, and/or special nuclear material	7. Chemical and/or physical form	8. Maximum amount that licensee may possess at any one time under this license
A. Phosphorus 32	A. As identified in 35.300	A. 200 millicuries
B. Strontium 89	B. As identified in 35.300	B. 200 millicuries
C. Iodine 131	C. As identified in 35.300	C. 200 millicuries
D. Strontium 90	D. As identified in 35.400	D. 115 millicuries
E. Iodine 125	E. As identified in 35.400	E. 1000 millicuries

(b)(7)(F)

9. Authorized use:
- A. through C. Any radiopharmaceutical therapy procedure approved in 10 CFR 35.300 for which the patient can be released under the provisions of 10 CFR 35.75.
  - D. Treatment of superficial eye conditions using an applicator distributed pursuant to 10 CFR 32.74.
  - E. Any brachytherapy procedure approved in 10 CFR 35.400 for which the patient can be released under the provisions of 10 CFR 35.75.

(b)(7)(F)

**MATERIALS LICENSE  
SUPPLEMENTARY SHEET**

License Number  
29-02023-06

Docket or Reference Number  
030-32352

Amendment No. 11

(b)(7)(F)

**CONDITIONS**

- 10. A. Licensed material may be used only at the licensee's facilities at 785 Totowa Road, Totowa, New Jersey, and at the Radiation Medicine Center at Cape May, 410 Route 9 North, Cape May Court House, New Jersey.
- B. Licensed material in items 6.E. may be used at the Clifton Surgery Center, 1117 Route 46 East, Clifton, New Jersey, and at Wills Surgery Center of Cape May County, 106 Courthouse S. Dennisville Road, Cape May Court House, New Jersey.
- 11. The Radiation Safety Officer for this license is Sam I. Brown, M.D.
- 12. Licensed material listed in Item 6 above is only authorized for use by, or under the supervision of, the following individuals for the materials and uses indicated:

Authorized Users

Sam I. Brown, M.D.

Material and Use

Phosphorus 32 as identified in 10 CFR 35.300  
Strontium 89 as identified in 10 CFR 35.300  
Iodine 131 as identified in 10 CFR 35.300  
Strontium 90 as identified in 10 CFR 35.400  
Iodine 125 as identified in 10 CFR 35.400

(b)(7)(F)

Arthur Harvey, M.D.

Phosphorus 32 as identified in 10 CFR 35.300  
Strontium 89 as identified in 10 CFR 35.300  
Iodine 131 as identified in 10 CFR 35.300  
Strontium 90 as identified in 10 CFR 35.400  
Iodine 125 as identified in 10 CFR 35.400

(b)(7)(F)

**MATERIALS LICENSE  
SUPPLEMENTARY SHEET**

License Number  
29-02023-06

Docket or Reference Number  
030-32352

Amendment No. 11

Authorized Users

Sung I. Lee, M.D.

Material and Use

Phosphorus 32 as identified in 10 CFR 35.300  
Strontium 89 as identified in 10 CFR 35.300  
Iodine 131 as identified in 10 CFR 35.300  
Strontium 90 as identified in 10 CFR 35.400  
Iodine 125 as identified in 10 CFR 35.400

(b)(7)(F)

John Leung, M.D.

Phosphorus 32 as identified in 10 CFR 35.300  
Strontium 89 as identified in 10 CFR 35.300  
Iodine 131 as identified in 10 CFR 35.300  
Strontium 90 as identified in 10 CFR 35.400  
Iodine 125 as identified in 10 CFR 35.400

(b)(7)(F)

13. The Medical Physicists for this license are Sreenivasa Murthy, M.S., Xuan Gen Chen, Ph.D., Michael P. Nunno, Robert Stanton, and Cheryl Trail.
14. In addition to the possession limits in Item 8, the licensee shall further restrict the possession of licensed material to quantities below the minimum limit specified in 10 CFR 30.35(d), 40.36(b), and 70.25(d) for establishing financial assurance for decommissioning.
15. In lieu of 10 CFR 35.404(a), immediately after retracting the source from the patient into its shielded position in the remote afterloading device, radiation survey shall be made of the patient and the remote afterloading device with a portable radiation detection survey instrument to confirm that the source has been removed from the patient. Records of the survey shall be maintained in lieu of the record required in 10 CFR 35.404(b).

**MATERIALS LICENSE  
SUPPLEMENTARY SHEET**

License Number  
29-02023-06

Docket or Reference Number  
030-32352

Amendment No. 11

16. In lieu of the source inventory described in 10 CFR 35.406, the licensee shall:
- A. Promptly determine that all sources have returned to the safe, shielded position at the conclusion of each high dose remote brachytherapy procedure.
  - B. Promptly make a survey of the area of use to confirm that no sources have been misplaced.
  - C. Make a record of the survey including survey instrument used, dose rate, time, date and name of the individual making the survey.
  - D. Retain the record of the survey in lieu of the record required in 10 CFR 35.406(d).

(b)(7)(F)

**MATERIALS LICENSE  
SUPPLEMENTARY SHEET**License Number  
29-02023-06Docket or Reference Number  
030-32352

Amendment No. 11

(b)(7)(F)

20. The licensee is authorized to transport licensed material in accordance with the provisions of 10 CFR Part 71, "Packaging and Transportation of Radioactive Material."



**MATERIALS LICENSE  
SUPPLEMENTARY SHEET**

License Number  
29-02023-06

Docket or Reference Number  
030-32352

Amendment No. 11

21. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below. The Nuclear Regulatory Commission's regulations shall govern unless the statements, representations and procedures in the licensee's application and correspondence are more restrictive than the regulations.

- A. Application dated April 23, 1991
- B. Letter dated July 8, 1991
- C. Letter dated July 26, 1991
- D. Letter dated August 5, 1993
- E. Letter dated February 3, 1994
- F. Letter dated March 24, 1994
- G. Letter dated May 25, 1994
- H. Letter dated January 22, 1996
- I. Letter dated March 1, 1996
- J. Letter dated September 12, 1996
- K. Letter dated September 26, 1996
- L. Letter dated November 4, 1996
- M. Letter dated January 7, 1997
- N. Letter dated January 16, 1997
- O. Letter dated January 30, 1997
- P. Letter dated November 25, 1997 except Quality Management Program
- Q. Letter dated March 16, 1999
- R. Letter dated April 7, 1999
- S. Letter dated May 20, 1999
- T. Letter dated February 22, 2000
- U. Letter dated April 11, 2000
- V. Letter dated May 22, 2000

For the U.S. Nuclear Regulatory Commission

Date May 25, 2000

By

*Original signed by Michelle Beardsley*

Michelle Beardsley  
Nuclear Materials Safety Branch 1  
Division of Nuclear Materials Safety  
Region I  
King of Prussia, Pennsylvania 19406