

From: [Gaskins, Farrah](mailto:Gaskins.Farrah)
To: Frank.Gantek@STMH.org
Subject: Request for additional information for license no. 06-00679-01
Date: Tuesday, November 24, 2015 3:50:00 PM

Dear Mr. Gantek,

Please send a return email to confirm that you have received this message.

This is in reference to the application dated October 19, 2015 requesting to renew Nuclear Regulatory Commission License No. 06-00679-01. In order to complete your renewal application, we need the following additional information:

1. You requested authorization to use materials under 35.400. According to the last inspection conducted December 16, 2014, materials authorized by 35.400 had not been used at your facility since 2011. Please confirm if you wish to maintain this authorization on your license.
2. Your current license authorizes materials under 35.200 to be used at 160 Robins Street, Waterbury CT. Your application is not specific about what material authorizations you are requesting for that location. Please confirm if only 35.200 will be used at this location.
3. Describe transport of materials from the Nuclear Medicine Suite the Cardiology department on the 2nd floor of the hospital.
4. Please describe the emergency response equipment available for manual brachytherapy facilities. This equipment should include gloves, reverse action tweezers, shielded containers, a low energy gamma scintillation survey instrument, and caution radioactive materials labels.
5. Please confirm if the request for 25.200 will include the use of Positron Emission Tomography (PET) radiopharmaceuticals. If so please provide shielding calculations and a description of the additional facilities and equipment for these uses.

Current NRC regulations and guidance are included on the NRC's website at www.nrc.gov; select Nuclear Materials; Med, In, & Academic Uses; then Licensee Toolkits, see our toolkit index page. You may also obtain these documents by contacting the Government Printing Office (GPO) toll-free at 1-866-512-1800. The GPO is open from 8:00 a.m. to 5:30 p.m. EST, Monday through Friday (except Federal holidays).

You may respond to my attention in writing by letter, email (if scanned into a pdf format), or fax (610-337-5269), referencing mail control number 589184. Please ensure that your response is signed by an appropriate licensee management representative.

Please feel free to contact me with any questions you may have.

Farrak C. Gaskins

Health Physicist

U.S. NRC, Region 1

2100 Renaissance Blvd, Suite 100

King of Prussia, PA 19406

(610) 337-5143 office

(610) 337-5269 fax