

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

<p>1. LICENSEE/LOCATION INSPECTED:</p> <p>Toledo Cardiology Consultants, Inc. 2409 Cherry Street, Suite 100 Toledo, OH 43608 Location Inspected: Monroe, MI</p> <p>REPORT NUMBER(S) 2015-002</p>	<p>2. NRC/REGIONAL OFFICE</p> <p>Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352</p>	
<p>3. DOCKET NUMBER(S)</p> <p>150-00034</p>	<p>4. LICENSE NUMBER(S)</p> <p>OH-02220490000</p>	<p>5. DATE(S) OF INSPECTION</p> <p>DECEMBER 7, 2015</p>

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:



- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

Non-cited violation(s) were discussed involving the following requirement(s):

- 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
(Violations and Corrective Actions)

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE	DATE
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Ryan Craffey		12/7/15
BRANCH CHIEF	Aaron McCraw		12/8/15

Docket File Information

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

<p>1. LICENSEE/LOCATION INSPECTED:</p> <p>Toledo Cardiology Consultants, Inc. 2409 Cherry Street, Suite 100 Toledo, OH 43608 Location Inspected: Monroe, MI</p> <p>REPORT NUMBER(S) 2015-002</p>	<p>2. NRC/REGIONAL OFFICE</p> <p>Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352</p>
--	---

<p>3. DOCKET NUMBER(S)</p> <p>150-00034</p>	<p>4. LICENSE NUMBER(S)</p> <p>OH-02220490000</p>	<p>5. DATE(S) OF INSPECTION</p> <p>December 7, 2015</p>
---	---	---

<p>6. INSPECTION PROCEDURES USED</p> <p>87130</p>	<p>7. INSPECTION FOCUS AREAS</p> <p>All</p>
---	---

SUPPLEMENTAL INSPECTION INFORMATION

<p>1. PROGRAM CODE(S)</p> <p>02220</p>	<p>2. PRIORITY</p> <p>3</p>	<p>3. LICENSEE CONTACT</p> <p>Tim Depinet, CNMT</p>	<p>4. TELEPHONE NUMBER</p> <p>(419) 251-3700</p>
--	-----------------------------	---	--

Main Office Inspection Next Inspection Date: TBD

Field Office Inspection _____

Temporary Job Site Inspection 1094 North Monroe Street, Monroe, MI

PROGRAM SCOPE

This was an announced field inspection of a cardiology clinic authorized by its State of Ohio license to perform cardiovascular imaging and localization studies at three facilities and at temporary job sites in Ohio jurisdiction. The licensee conducted these activities at temporary job sites in NRC jurisdiction under a reciprocity request filed in December 2014 for calendar year 2015. At the time of the inspection, the licensee's only client in NRC jurisdiction was a cardiologist in Monroe, Michigan. The licensee received unit doses for this work at one of its facilities in Ohio, and transported the doses along with an imaging camera, dose calibrator with well counter, survey meter and other equipment to the client's location for use. The licensee returned all waste generated at the client's facility to one of the authorized facilities in Ohio for decay in storage.

PERFORMANCE OBSERVATIONS

The inspector toured the client's facility in Monroe, Michigan to evaluate the licensee's implementation of measures for material security, hazard communication and exposure control. The inspector conducted independent and confirmatory surveys of the facility, and found no exposures to members of the public in excess of regulatory limits. The inspector observed the preparation and administration of two cardiac stress tests using technetium-99m. The attending technologist also demonstrated the implementation of licensee procedures for area surveys, spill response and waste handling. The technologist also demonstrated the satisfactory use of dosimetry, survey instruments, and ALARA practices, and through various discussions demonstrated an adequate understanding of radiation protection principles and applicable regulatory requirements.

The inspector reviewed a selection of the licensee's records available on-site, including shipping papers, dosimetry reports, and quarterly audits conducted by the licensee's medical physics consultant of the mobile nuclear medicine service.

No violations of NRC requirements were identified as a result of this inspection.