



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION IV
1600 EAST LAMAR BLVD
ARLINGTON, TEXAS 76011-4511

EMAIL



Name: Rex G. Hoy License: 25-11498-01
Docket: 030-05179
Control: 588800
Organization: State of Montana, Department of Transportation
Phone: 406-444-6270
E-mail Address: rhoy@mt.gov
From: Jacqueline D. Cook
Date: December 11, 2015
Subject: Application dated September 10, 2015 for License Amendment
Pages: 2

Mr. Hoy:

Per your application dated September 10, 2015, the items on the next page are deficiencies which require your response. **Please respond to this e-mail at your earliest convenience but by no later than COB Monday, January 4, 2016.** Our fax number is (817) 200-1263. You may respond by e-mail in pdf format if you'd like. My email address is Jackie.Cook@nrc.gov. When responding to this e-mail, please include the license, docket and control numbers located at the top of this page.

Please note that your renewal application was reviewed using NUREG-1556, Vol. 1, Rev. 1, "Consolidated Guidance About Materials Licenses: Program-Specific Guidance About Portable Gauge Licenses" dated November 2001 (<http://www.nrc.gov/reading-rm/doc-collections/nuregs/staff/sr1556/v1/r1/>).

Thanking you in advance for your cooperation, assistance, and prompt response in this matter.

/RA/
Jacqueline D. Cook
Senior Health Physicist

PUBLIC

- Immediate Release
 Normal Release

NON-PUBLIC

- Health, Safety, or Environment Related
 Confidential
 Other

Reviewed: *JDC* Date: *12/11/15*

1. For non-routine maintenance or repair operations that require detaching the source or source rod from the gauge, you stated that you had the experience listed in NUREG-1556, Vol. 1, Rev. 2, Appendix F. Please note that NUREG-1556, Vol. 1, Rev. 2 has not been finalized yet.

In addition, please note that Appendix F of the current NUREG-1556, Vol. 1, Rev. 1 is the portable gauge audit checklist. Appendix G is the information needed to support applicant's request to perform non-routine maintenance. Please provide the applicable information from Appendix G for you to be approved to perform non-routine maintenance or repair operations requiring detaching of the source or source rod from the portable gauge, including your training and experience.

PUBLIC

- Immediate Release
- Normal Release

NON-PUBLIC

- A.3 Sensitive-Security Related
- A.7 Sensitive Internal
- Other: _____

Reviewer: _____

Date: _____

[Handwritten Signature] *12/1/15*