

Lawyer, Dennis

From: Fendt, William A. <wfendt@udel.edu>
Sent: Thursday, December 10, 2015 2:43 PM
To: Lawyer, Dennis
Subject: [External_Sender] RE: NRC Renewal

Mr. Lawyer-

The 51mCi limit for the Po-210 sources is good.

Addresses:

Hugh R. Sharp Campus at 1044 College Drive, Lewes, DE 19958

Georgetown Campus at 16483 County Seat Hwy, Georgetown, DE 19947

Bill

William Fendt
Radiation and Laser Safety Officer
University of Delaware
Environmental Health & Safety
Newark, DE
wfendt@udel.edu

From: Lawyer, Dennis [mailto:Dennis.Lawyer@nrc.gov]
Sent: Thursday, December 10, 2015 2:33 PM
To: Fendt, William A. <wfendt@udel.edu>
Subject: NRC Renewal

Mr. Fendt,

Sorry for the extra communications.

I changed your maximum for Po-210 to 51 mCi because the maximum per source is really 10.2 mCi according the sealed source and device registry so to ensure you can have 5 units, I made it 51 mCi. I hope that is satisfactory to you.

Second, it was recently decided to change how we write location of use for a campus. For example, your Newark Campus will be listed as Newark Campus at 222 South Chapel Street, Newark, Delaware. Could you provide me a specific address to use for your other campuses (Hugh R. Sharp, and Georgetown)? You can just email this information back to me.

Dennis Lawyer
Health Physicist
U.S. Nuclear Regulatory Commission
Division of Nuclear Material Safety
610-337-5366
610-337-5269 (F)