



Architecture • Civil Engineering • Land Surveying • Site Development • Geotechnical Engineering • Inspection & Materials Testing

FACSIMILE TRANSMITTAL SHEET

TO: JENNIFER BISHOP 12-16-2015
Name Date

NRC
Company

630 515 1078
Fax Number

FROM: THOMAS G. BEULKE
Name

RSO
Project Number/Description

PAGES (INCLUDING COVER SHEET): 2

NOTES/COMMENTS:

CONTROL # 588316

IF YOU HAVE NOT RECEIVED ALL PAGES SENT, OR IF DATA IS UNCLEAR, PLEASE CALL 636-584-0540. THANK YOU.

8 East Main Street
Wentzville, Missouri 63385
Telephone: 636-332-4574
Fax: 636-327-0760

530A East Independence Drive
Union, Missouri 63084
Telephone: 636-584-0540
Fax: 636-584-0512
www.cochraneng.com

737 Rudder Road
Fenton, Missouri 63026
Telephone: 314-842-4033
Fax: 314-842-5957

RECEIVED DEC 16 2015



Architecture • Civil Engineering • Land Surveying • Site Development • Geotechnical Engineering • Inspection & Materials Testing

December 31, 2016

Attn: Jennifer Bishop

Materials Licensing Branch

U.S. Nuclear Regulatory Commission, Region III

2443 Warrenville Road, Suite 210

Lisle, IL 60532-4352

RE: Location Close-out for Storage of Nuclear Gauge
License No. 24-32594-01
Docket No. 030-37025

Ms. Bishop,

As requested, we are sending this letter as a follow-up to our submission, by fax, of our final close-out inspection at 8 West Main Street, Wentzville, MO and Leak Test performed on 12-4-15. At this time, this location no longer contains radioactive materials and all inspections have been completed.

Sincerely,

Thomas G. Beuke

Radiation Safety Officer

Cochran

8 East Main Street
Wentzville, Missouri 63385
Telephone: 636-332-4574
Fax: 636-327-0760

530A East Independence Drive
Union, Missouri 63084
Telephone: 636-584-0540
Fax: 636-584-0512

44 Camden Court SE/PO Box 1138
Camdenton, MO 65020
Telephone: 573-525-0299
Fax: 573-525-0298

737 Rudder Road
Fenton, Missouri 63026
Telephone: 314-842-4033
Fax: 314-842-5957

www.cochraneng.com

LEAK TEST CERTIFICATE

Wipe C 12-4-15

DESCRIPTION OF DEVICE/SOURCE

Model HS-500/EZ Ser 8419
SOURCE 1 SOURCE 2
Material Cs-137 Am-241:Be
Source Ser 4574C9 330-13
RSO Name Thomas G. Beuke
Telephone (636) 221-1986

Do Not Write In This Space
For HSI Use Only

REMOVABLE ACTIVITY
Bet/Gam Alpha
0 uCi 0 uCi
Humboldt Scientific
By Allen Parker
Date Dec 15

Instructions

- 1 Use wipe procedures as described in the device manual.
- 2 Enter all information under description.
- 3 Print or type return address in the space provided.
- 4 Wipe source (s) and put filter paper in the plastic bag.
- 5 Keep the middle copy and mail this form and bag to Humboldt.
- 6 Regulations require that sources with removable activity greater than 0.005 uCi be removed from service for repair and decontamination or disposal. Authorities must be notified.

5200174
Return
Address
Label
Please
Type or
Print
Clearly

- Cochran
- 530 A E. Independence
- Union, MO 63084

HUMBOLDT SCIENTIFIC INC. 2525 Atlantic Ave., Raleigh, NC 27604, (919) 832-3190

LEAK TEST CERTIFICATE

Wipe C 12-4-15

DESCRIPTION OF DEVICE/SOURCE

Model HS-600/EZ Ser 5149
SOURCE 1 SOURCE 2
Material Cs-137 Am-241:Be
Source Ser 8218Cm 476913
RSO Name Thomas G. Beuke
Telephone (636) 221-1986

Do Not Write In This Space
For HSI Use Only

REMOVABLE ACTIVITY
Bet/Gam Alpha
0 uCi 0 uCi
Humboldt Scientific
By Allen Parker
Date Dec 15

Instructions

- 1 Use wipe procedures as described in the device manual.
- 2 Enter all information under description.
- 3 Print or type return address in the space provided.
- 4 Wipe source (s) and put filter paper in the plastic bag.
- 5 Keep the middle copy and mail this form and bag to Humboldt.
- 6 Regulations require that sources with removable activity greater than 0.005 uCi be removed from service for repair and decontamination or disposal. Authorities must be notified.

5200174
Return
Address
Label
Please
Type or
Print
Clearly

- Cochran
- 530 A E. Independence
- Union, MO 63084

HUMBOLDT SCIENTIFIC INC. 2525 Atlantic Ave., Raleigh, NC 27604, (919) 832-3190

RECEIVED DEC 16 2015