

Part 21 (PAR)

Event # 51621

<b>Rep Org:</b> HOMEWOOD PRODUCTS CORPORATION		<b>Notification Date / Time:</b> 12/23/2015 11:15 (EST)	
<b>Supplier:</b> NATIONAL TECHNICAL SYSTEMS, INC. FORM		<b>Event Date / Time:</b> 12/21/2015 (EST)	
<b>Last Modification:</b> 12/23/2015			
<b>Region:</b> 1	<b>Docket #:</b>		
<b>City:</b> PITTSBURGH	<b>Agreement State:</b>		Yes
<b>County:</b>	<b>License #:</b>		
<b>State:</b> PA			
<b>NRC Notified by:</b> RICHARD MARTIN		<b>Notifications:</b> JAMES NOGGLE R1DO	
<b>HQ Ops Officer:</b> HOWIE CROUCH		PART 21/50.55 REACTORS EMAIL	
<b>Emergency Class:</b> NON EMERGENCY			
<b>10 CFR Section:</b>			
21.21(d)(3)(i)		DEFECTS AND NONCOMPLIANCE	

**PART 21 REPORT - ERROR IN REPORTING THE VARIABILITY IN THE DOSE DELIVERED ON CERTIFICATE OF PROCESSING**

The following information was received from Homewood Products Corporation via fax:

"Steris Isomedix Services measuring and test equipment used by NTS (Formerly Wyle Laboratories, Inc.) to determine the applied radiation dose did not account for all of the uncertainties involved, and therefore the actual radiation dose applied to components and reported to Homewood Products Corporation by Wyle Laboratories, could be less than the requested service condition dose."

The following information was received from Homewood Products Corporation as provided to them by National Technical Services, Inc. (formerly Wyle Laboratories, Inc.):

"The defect is an error in reporting the variability in the dose delivered or lack thereof on the Isomedix Certificate of Processing for the Whippany, NJ facility. The U.S. Nuclear Regulatory Commission (NRC) under 10 CFR Part 50, Appendix B issued a Notice of Nonconformance 99901145/2014-201-01 to Steris stating that the measuring and testing equipment used to determine the applied radiation dose reported on the Isomedix Certificate of Processing provided with each gamma irradiation run did not account for all the uncertainties involved (i.e. density of unrelated products in carriers, off-carrier locations within the irradiator, and Cobalt-60 source decay) and therefore the actual radiation dose applied to components could be less than requested as reported on the Certificate of Processing.

"Steris Isomedix Services completed an evaluation of the dose rate variability of items processed in off-carrier locations in the irradiator. Steris Isomedix Services Position Paper dated 4/27/15 states the overall variability (uncertainty) associated with gamma radiation exposures at their Whippany, NJ facility."

IE19  
NM55

The affected facilities have been notified of the non-conformance. They are:

Homewood Products Corporation, Pittsburgh, PA  
Homewood Energy Services, Pittsburgh, PA

\*\*\*\*\*

# Homewood Products Corporation



## PART 21 MATERIAL REVIEW BOARD EVALUATION FORM

Date of Discovery: 12/21/15	Referenced Deficiency Document: Attached NTS Letters
-----------------------------	--

**Description of Problem:** Steris Isomedix Services measuring and test equipment used by NTS (Formerly Wyle Laboratories, Inc.) to determine the applied radiation dose did not account for all of the uncertainties involved, and therefore the actual radiation dose applied to components and reported to Homewood Products Corporation by Wyle Laboratories, could be less than the requested service condition dose.

Section 1: Screening:	Yes	No
<p>Is the condition a failure to comply or a deviation i.e.,</p> <ul style="list-style-type: none"> <li>a departure from the technical requirements included in a procurement document?</li> <li>a departure from standard design certification or standard design approval?</li> </ul> <p><b>Explain:</b> The condition is a departure from the technical requirements in a procurement document and design certification.</p>	X	<input type="checkbox"/>

Section 2: Capability	Yes	No
<p>Does the HPC have the capability to perform the evaluation to determine if a defect exists?</p> <p><i>If no, the HPC Material Review Board Chairman shall inform the purchasers or affected licensees within five (5) working days so that the purchasers or affected licensees may evaluate the deviation or failure to comply.</i></p>	X	<input type="checkbox"/>

Section 3: Evaluation	Yes	No
<p>Is the condition a defect or failure to comply that results in a substantial safety hazard i.e.,</p> <ul style="list-style-type: none"> <li>a significant deficiency in a final design?</li> <li>a significant deviation from equipment performance specifications or qualifications?</li> <li>a major degradation in performance of essential safety-related equipment?</li> <li>a significant breakdown in any portion of the Quality Assurance Program?</li> </ul> <p><b>Explain:</b> Homewood Products Corporation will perform an evaluation to determine the impact on the five (5) motor qualification reports supplied to Homewood Products, and if the results are a substantial safety hazard. The evaluation will be completed within the 60 day period from the above discovery date.</p>	X	<input type="checkbox"/>

Section 4: Notifications Required	Yes	Date
HPC Quality Manager	X	12/21/15
Nuclear Regulatory Commission (NRC)	X	12/23/15
Customer		

Additional actions to be taken: (references)

HPC Attendee Review	Organization	HPC Attendee Review	Organization
Richard J. Martin	Quality Assurance	John Mikach	Engineering
Kevin Monaco	HES Marketing		

HPC Material Review Board Chair Approval Signature: *Gary DeGroot* Date: 12/23/2015



Formerly Wyle Laboratories, Inc.

December 14, 2015

**Homewood Products Corporation**  
**820 Washington Blvd.**  
**Pittsburgh, PA 15206**

**Attention: QA Manager**

**Subject:** Pursuant to 10 CFR Part 21, this letter notifies Homewood Products of the existence of a reportable defect and its evaluation

**Reference:** Homewood Products Purchase Order No. N43333, Wyle Laboratories Job Number 55622

The defect is an error in reporting the variability in the dose delivered or lack thereof on the Isomedix Certificate of Processing for the Whippany, NJ facility. The U.S. Nuclear Regulatory Commission (NRC) under 10 CFR Part 50, Appendix B issued a Notice of Nonconformance 99901145/2014-201-01 to Steris stating that the measuring and testing equipment used to determine the applied radiation dose reported on the Isomedix Certificate of Processing provided with each gamma irradiation run did not account for all the uncertainties involved (i.e. density of unrelated products in carriers, off-carrier locations within the irradiator, and Cobalt-60 source decay) and therefore the actual radiation dose applied to components could be less than requested as reported on the Certificate of Processing.

Steris Isomedix Services completed an evaluation of the dose rate variability of items processed in off-carrier locations in the irradiator. Steris Isomedix Services Position Paper dated 4/27/15 states the overall variability (uncertainty) associated with gamma radiation exposures at their Whippany NJ facility.

Steris provided revised Component Irradiation Certification(s) that shows a revised minimum total delivered dose of 17.22 Mrad for Steris Process Run ID 30224A. The revised Steris certificate is attached.

Per the Wyle Laboratories Report 55622TR09, the stated required service condition dose is 18.25 Mrad.

The revised minimum total delivered dose does not envelop the stated service condition. The impact of this Part 21 on qualification is unknown. Please review this information to determine the impact of this defect on your application.



*Formerly Wyle Laboratories, Inc.*

December 14, 2015

If you have any questions, please feel free to contact me by phone (256) 716-4276, fax (256) 830-2109, or email [megan.toomey@nts.com](mailto:megan.toomey@nts.com).

Sincerely,

**National Technical Systems Huntsville**  
Formerly Wyle Laboratories, Inc.

A handwritten signature in black ink, appearing to read "Megan Toomey", with a long horizontal flourish extending to the right.

Megan Toomey  
Senior Contracts Manager



Formerly Wyle Laboratories, Inc.

December 14, 2015



**COMPONENT IRRADIATION CERTIFICATION**

Prepared for Wyle Laboratories

Source Type: Cobalt 60 Gamma P.O. HSV0049082

Air Equip. Required Dose (MRADS)		12.00			
Rate Not to Exceed (MRADS / Hr.)		1.00			
<b>SPECIMENS:</b>					
Qty	Part No.	Serial No.	Description		
1	N/A	N/A	75 hp Horizontal Motor		
1	N/A	N/A	Chevron SRI-I Grease		
<b>DATA:</b>					
Total Delivered Dose (Air) MRADS:		Min	17.22	Max	22.58
Dose Rate (Air) MRADS / Hr:		Min	.80	Max	.88
Total Exposure Hours:		23.76			
Spec Rotation	2-WAY:	Turntable Rotation	None:		
Date In:	08/11/09	Date Out:	08/13/09		
<b>DOSIMETRY:</b>					
Dosimeter Type	Harwell Rad 4034 Perspex	Batch	KC		
Calibration Date	01/20/09	Calibration Due Date	01/20/10		
Readout Instrument:	Beckman DU-640	Serial No.:	4328387		
Calibration Date	08/15/08	Calibration Due Date	08/18/09		
Comments: Process Run #30224A / "Dose has been adjusted for a process variability of 9.4% in accordance with STERIS Isomedix Services Position Paper Appendix C (Rev 4/27/15)" / Revised Copy 11/01/15					
<b>ATTACHMENTS:</b>					
Worksheets:	N/A	Drawings:	N/A		
Notice of Anomaly:	N/A				

Approved By: *[Signature]* Title: Sr. QS/RC Analyst Date: 12/03/15

Processing Location: STERIS Isomedix Services 9 Apollo Drive, Whippany, NJ 07981 Phone: 973-887-2754 Fax: 973-887-5581  
 The product run described above was processed in accordance with STERIS Isomedix Services Quality System requirements and the approved process plan for STERIS Isomedix Services facilities in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA). STERIS Isomedix Services operates under a quality system which meets the requirements of the FDA QSR and ISO 13485:2003. STERIS Isomedix Services adheres to regulatory provisions in ANSI/AAMI/ISO 11177 and ISO 9001.



Formerly Wyle Laboratories, Inc.

December 14, 2015

**Homewood Energy Services**  
820 Washington Blvd.  
Pittsburgh, PA 15206

**Attention: QA Manager**

**Subject:** Pursuant to 10 CFR Part 21, this letter notifies Homewood Energy Services of the existence of a reportable defect and its evaluation

**Reference:** Homewood Energy Services Purchase Order No. N37223, Wyle Laboratories Job Number 51526

The defect is an error in reporting the variability in the dose delivered or lack thereof on the Isomedix Certificate of Processing for the Whippany, NJ facility. The U.S. Nuclear Regulatory Commission (NRC) under 10 CFR Part 50, Appendix B issued a Notice of Nonconformance 99901145/2014-201-01 to Steris stating that the measuring and testing equipment used to determine the applied radiation dose reported on the Isomedix Certificate of Processing provided with each gamma irradiation run did not account for all the uncertainties involved (i.e. density of unrelated products in carriers, off-carrier locations within the irradiator, and Cobalt-60 source decay) and therefore the actual radiation dose applied to components could be less than requested as reported on the Certificate of Processing.

Steris Isomedix Services completed an evaluation of the dose rate variability of items processed in off-carrier locations in the irradiator. Steris Isomedix Services Position Paper dated 4/27/15 states the overall variability (uncertainty) associated with gamma radiation exposures at their Whippany NJ facility.

Steris provided revised Component Irradiation Certification(s) that shows a revised minimum total delivered dose of 8.73 Mrad for Steris Process Run ID 29620. The revised Steris certificate is attached.

Per the Wyle Laboratories Report 51526TR05, the stated required service condition dose is 8.75 Mrad.

The revised minimum total delivered dose does not envelop the stated service condition. The impact of this Part 21 on qualification is unknown. Please review this information to determine the impact of this defect on your application.

If you have any questions, please feel free to contact me by phone (256) 716-4276, fax (256) 830-2109, or email [megan.toomey@nts.com](mailto:megan.toomey@nts.com).

Sincerely,

**National Technical Systems Huntsville**  
Formerly Wyle Laboratories, Inc.

A handwritten signature in black ink, appearing to read "Megan Toomey", is written over a horizontal line.

Megan Toomey  
Senior Contracts Manager



Formerly Wyle Laboratories, Inc.

December 14, 2015

STERIS-Isomedix Services  
 9 Apollo Drive  
 Whippany NJ 07981

STERIS-Isomedix Services Component Irradiation Certification					
CUSTOMER	Wyle Laboratories		P. O. #	HSV0034340	
AIR EQUIV. REQUIRED DOSE (MRADS)			9.63		
RATE NOT TO EXCEED (MRADS/HR)			1.00		
SPECIMENS:					
QTY	PART NO.	SERIAL NO.	DESCRIPTION		
1	N/A	N/A	5 HP Motor		
1	N/A	N/A	4 kv Motorette		
1	N/A	N/A	Oil (3/4 full)		
DATA:					
SOURCE TYPE: COBALT-60 GAMMA					
TOTAL DELIVERED DOSE (AIR) MRADS:		MIN:	8.73	MAX:	12.37
DOSE RATE (AIR) MRADS/HR:		MIN:	0.81	MAX:	0.95
TOTAL EXPOSURE HOURS:		11.9			
SPECIMEN ROTATION		2-WAY:	X	4-WAY:	NONE
DATE IN:		03/08/05		DATE OUT: 03/09/05	
DOSIMETRY:					
DOSIMETER TYPE:	HARWELL PERSPEX	BATCH:	NK		
	±6.5%	CALIBRATION DATE	10/21/04		
READOUT INSTRUMENT:		Beckman DU-640-C			
SERIAL NO.:	4324039	CALIBRATION DATE	09/14/04		
COMMENTS:	Irradiation Control #29620 / *Dose has been adjusted for a process variability of 9.4% in accordance with STERIS Isomedix Services Position Paper Appendix C (Rev 4/27/15) / Revised Copy 11/06/15				
ATTACHMENTS:					
WORKSHEETS:	N/A		DRAWINGS:	N/A	
NOTICE OF ANOMOLY:		N/A			
AUTHORIZED SIGNATURE:					
TITLE:	Sr. OS/RC Analyst		DATE:	11/06/15	

Form 1700.3N1  
 Revision: A  
 Effective Date: 09/18/03



Formerly Wyle Laboratories, Inc.

December 14, 2015

**Homewood Products Corporation  
820 Washington Blvd.  
Pittsburgh, PA 15206**

**Attention: QA Manager**

**Subject:** Pursuant to 10 CFR Part 21, this letter notifies Homewood Products of the existence of a reportable defect and its evaluation.

**Reference:** Homewood Products Purchase Order No. N42313, Wyle Laboratories Job Number 54067

The defect is an error in reporting the variability in the dose delivered or lack thereof on the Isomedix Certificate of Processing for the Whippany, NJ facility. The U.S. Nuclear Regulatory Commission (NRC) under 10 CFR Part 50, Appendix B issued a Notice of Nonconformance 99901145/2014-201-01 to Steris stating that the measuring and testing equipment used to determine the applied radiation dose reported on the Isomedix Certificate of Processing provided with each gamma irradiation run did not account for all the uncertainties involved (i.e. density of unrelated products in carriers, off-carrier locations within the irradiator, and Cobalt-60 source decay) and therefore the actual radiation dose applied to components could be less than requested as reported on the Certificate of Processing.

Steris Isomedix Services completed an evaluation of the dose rate variability of items processed in off-carrier locations in the irradiator. Steris Isomedix Services Position Paper dated 4/27/15 states the overall variability (uncertainty) associated with gamma radiation exposures at their Whippany NJ facility.

Steris provided revised Component Irradiation Certification(s) that shows a revised minimum total delivered dose of:

181.2 Mrad for Steris Process Run ID 27595.

176.4 Mrad for Steris Process Run ID 27608.

The revised Steris certificates are attached.

Per the Wyle Laboratories Report 54067TR08, the stated required service condition dose is 200.0 Mrad.

The revised minimum total delivered dose does not envelop the stated service condition. The impact of this Part 21 on qualification is unknown. Please review this information to determine the impact of this defect on your application.



*Formerly Wyle Laboratories, Inc.*

December 14, 2015

If you have any questions, please feel free to contact me by phone (256) 716-4276, fax (256) 830-2109, or email [megan.toomey@nts.com](mailto:megan.toomey@nts.com).

Sincerely,

**National Technical Systems Huntsville**  
Formerly Wyle Laboratories, Inc.

A handwritten signature in black ink, appearing to read "Megan Toomey", with a long horizontal flourish extending to the right.

Megan Toomey  
Senior Contracts Manager





Formerly Wyle Laboratories, Inc.

December 14, 2015



**COMPONENT IRRADIATION CERTIFICATION**

Prepared for Wyle Laboratories

Source Type: Cobalt 60 Gamma P.O. #HSV0043031

Air Eqvty. Required Dose (MRADS)		200.00	
Rate Not to Exceed (MRADS / Hr.)		1.00	
<b>SPECIMENS:</b>			
Qty	Part No.	Serial No.	Description
3	Project ID: AWG Lead Wires	N/A	Cables (#4 gauge)
N/A	N/A	N/A	N/A
<b>DATA:</b>			
Total Delivered Dose (Air) MRADS		Min	176.40
		Max	243.96
Dose Rate (Air) MRADS / Hr		Min	0.60
		Max	0.66
Total Exposure Hours		133.34	
Static Rotation	2 - WAY	Turntable Rotation	None
Date In:	10/30/07	Date Out:	11/24/07
<b>DOSIMETRY:</b>			
Dosimeter Type	Harwell Red 4034 Perspex	Batch	M
Calibration Date	11/08/07	Calibration Due Date	11/08/08
Readout Instrument	Beckman DU-640	Serial No.	4326387
Calibration Date	08/28/07	Calibration Due Date	08/28/08
Comments: Process Run #27608 / Dose has been adjusted for a process variability of 11.8% in accordance with STERIS Isomatrix Services Position Paper Appendix C (Rev 4/27/15) / Revised Copy 11/10/15			
<b>ATTACHMENTS:</b>			
Worksheets	N/A	Drawings	N/A
Notice of Anomaly	N/A		

Approved By: [Signature] Title: Sr. QS/RC Analyst Date: 11/10/15

Processing Location: STERIS Isomatrix Services 49 Apollo Drive, Whippany, NJ 07981 Phone: 973-897-3754 Fax: 973-897-6991  
 STERIS Isomatrix Services facilities are in compliance with applicable state and federal regulations of the NRC, 10 CFR, and OSHA. STERIS Isomatrix Services operates under a quality system which meets the requirements of the FDA QSR and ISO 13485:2001. STERIS Isomatrix Services adheres to requirements provided through ANSI/ASME/ISO 11137 and 13520.

PROC-00830 Form 3 Rev: 1 Eff Date: Aug 8, 2007 Status: 07% Completed Page 1 of 1  
 Multiple or Single Facility



Formerly Wyle Laboratories, Inc.

December 14, 2015

**Homewood Energy Services**  
820 Washington Blvd.  
Pittsburgh, PA 15206

**Attention: QA Manager**

**Subject:** Pursuant to 10 CFR Part 21, this letter notifies Homewood Energy Services of the existence of a reportable defect and its evaluation.

**Reference:** Homewood Energy Services Purchase Order Nos. N35384 and N35662, Wyle Laboratories Job Number 48955

The defect is an error in reporting the variability in the dose delivered or lack thereof on the Isomedix Certificate of Processing for the Whippany, NJ facility. The U.S. Nuclear Regulatory Commission (NRC) under 10 CFR Part 50, Appendix B issued a Notice of Nonconformance 99901145/2014-201-01 to Steris stating that the measuring and testing equipment used to determine the applied radiation dose reported on the Isomedix Certificate of Processing provided with each gamma irradiation run did not account for all the uncertainties involved (i.e. density of unrelated products in carriers, off-carrier locations within the irradiator, and Cobalt-60 source decay) and therefore the actual radiation dose applied to components could be less than requested as reported on the Certificate of Processing.

Steris Isomedix Services completed an evaluation of the dose rate variability of items processed in off-carrier locations in the irradiator. Steris Isomedix Services Position Paper dated 4/27/15 states the overall variability (uncertainty) associated with gamma radiation exposures at their Whippany NJ facility.

Steris provided revised Component Irradiation Certification(s) that shows a revised minimum total delivered dose of:

- 99.66 Mrad for Steris Process Run ID 22503 (4.0 kV Motorette).
  - 99.66 Mrad for Steris Process Run ID 22503 (6.6 kV Motorette).
  - 19.93 Mrad for Steris Process Run ID 22503 (10 hp Motor/Grease).
- The revised Steris certificates are attached.

Per the Wyle Laboratories Report 48955TR04, the stated required service condition dose is either 100.0 Mrad or 20.0 Mrad.



*Formerly Wyle Laboratories, Inc.*

December 14, 2015

The revised minimum total delivered dose does not envelop the stated service condition. The impact of this Part 21 on qualification is unknown. Please review this information to determine the impact of this defect on your application.

If you have any questions, please feel free to contact me by phone (256) 716-4276, fax (256) 830-2109, or email [megan.toomey@nts.com](mailto:megan.toomey@nts.com).

Sincerely,

**National Technical Systems Huntsville**  
Formerly Wyle Laboratories, Inc.

A handwritten signature in black ink, appearing to read "Megan Toomey", with a long horizontal flourish extending to the right.

Megan Toomey  
Senior Contracts Manager



Formerly Wyle Laboratories, Inc.

December 14, 2015

STERIS-Isomedix Services  
 9 Apollo Drive  
 Whippany NJ 07981

STERIS-Isomedix Services Component Irradiation Certification			
CUSTOMER	Wyle Laboratories	P. O. #	HSV0031928
AIR EQUIV. REQUIRED DOSE (MRADS)	110.00		
RATE NOT TO EXCEED (MRADS/HR)	1.00		
SPECIMENS:			
QTY	PART NO.	SERIAL NO.	DESCRIPTION
1	N/A	N/A	4.0 kV Motorette
		N	
		A	
DATA:			
SOURCE TYPE: COBALT-60 GAMMA			
TOTAL DELIVERED DOSE (AIR) MRADS:	MIN:	99.66	MAX: 140.73
DOSE RATE (AIR) MRADS/HR:	MIN:	0.59	MAX: 0.63
TOTAL EXPOSURE HOURS:	185.44		
SPECIMEN ROTATION	2-WAY:	X	4-WAY: NONE:
DATE IN:	06/14/04	DATE OUT:	06/29/04
DOSIMETRY:			
DOSIMETER TYPE:	HARWELL PERSPEX	BATCH:	H8
	±6.5%	CALIBRATION DATE	02/04/04
READOUT INSTRUMENT:	Beckman DU-640-C		
SERIAL NO.:	4324039	CALIBRATION DATE	09/10/03
COMMENTS:	Irradiation Control Number 22503 / "Dose has been adjusted for a process variability of 9.4% in accordance with STERIS Isomedix Services Position Paper Appendix C (Rev 4/27/15)" / Revised Copy 11/06/15		
ATTACHMENTS:			
WORKSHEETS:	N/A	DRAWINGS:	N/A
NOTICE OF ANOMOLY:	N/A		
AUTHORIZED SIGNATURE:			
TITLE:	Sr. QS/RC Analyst	DATE:	11/06/15

Form 1702.3NJ  
 Revision: A  
 Effective Date: 08/18/03



Formerly Wyle Laboratories, Inc.

December 14, 2015

STERIS-Isomedix Services  
 9 Apollo Drive  
 Whippany NJ 07981

STERIS-Isomedix Services Component Irradiation Certification					
CUSTOMER	Wyle Laboratories		P. O. #	HSV0031928	
AIR EQUIV. REQUIRED DOSE (MRADS)		110.00			
RATE NOT TO EXCEED (MRADS/HR)		1.00			
SPECIMENS:					
QTY	PART NO.	SERIAL NO.	DESCRIPTION		
1	N/A	N/A	6.6 kV Motorotte		
		N			
		A			
DATA:					
SOURCE TYPE: COBALT-60 GAMMA					
TOTAL DELIVERED DOSE (AIR) MRADS:		MIN:	99.66	MAX:	138.39
DOSE RATE (AIR) MRADS/HR:		MIN:	0.60	MAX:	0.69
TOTAL EXPOSURE HOURS:		183.34			
SPECIMEN ROTATION		Z-WAY:	X	4-WAY:	NONE:
DATE IN:		06/14/04		DATE OUT:	06/29/04
DOSIMETRY:					
DOSIMETER TYPE:		HARWELL PERSPEX		BATCH:	HB
		±6.5%		CALIBRATION DATE	02/04/04
READOUT INSTRUMENT:		Beckman DU-640-C			
SERIAL NO.:		4324039		CALIBRATION DATE	09/10/03
COMMENTS:		Irradiation Control Number 22503 / "Dose has been adjusted for a process variability of 9.4% in accordance with STERIS Isomedix Services Position Paper Appendix C (Rev 4/27/15)" / Revised Copy 11/06/15			
ATTACHMENTS:					
WORKSHEETS:		N/A		DRAWINGS:	N/A
NOTICE OF ANOMOLY:		N/A			
AUTHORIZED SIGNATURE:					
TITLE:		Sr. QS/RC Analyst		DATE:	11/06/15

Form 1702.3N1  
 Revision: A  
 Effective Date: 08/18/03



Formerly Wyle Laboratories, Inc.

December 14, 2015

STERIS-Isomedix Services  
 9 Apollo Drive  
 Whippany NJ 07981

STERIS-Isomedix Services  
 Component Irradiation Certification

CUSTOMER	Wyle Laboratories		P. O. #	HSV0031928	
AIR EQUIV. REQUIRED DOSE (MRADS)	22.00				
RATE NOT TO EXCEED (MRADS/HR)	1.00				
SPECIMENS:					
QTY	PART NO.	SERIAL NO.	DESCRIPTION		
1	N/A	N/A	10 hp Motor		
1	N/A	N/A	Grease		
N/A	N/A	N/A	N/A		
DATA:					
SOURCE TYPE: COBALT-60 GAMMA					
TOTAL DELIVERED DOSE (AIR) MRADS:	MIN:	19.93	MAX:	27.88	
DOSE RATE (AIR) MRADS/HR:	MIN:	0.38	MAX:	0.44	
TOTAL EXPOSURE HOURS:	57.9				
SPECIMEN ROTATION	2-WAY:	X	4-WAY:	NONE:	
DATE IN:	06/14/14		DATE OUT:	06/18/04	
DOSIMETRY:					
DOSIMETER TYPE:	MARWELL PERSPEX	BATCH:	HB		
	±6.5%	CALIBRATION DATE	02/04/04		
READOUT (INSTRUMENT):	Beckman DU-640-C				
SERIAL NO.:	4324089	CALIBRATION DATE	09/10/03		
COMMENTS:	Irradiation Control Number 22503 / "Dose has been adjusted for a process variability of 9.4% in accordance with STERIS Isomedix Services Position Paper Appendix C (Rev 4/27/15)" / Revised Copy 11/06/15				
ATTACHMENTS:					
WORKSHEETS:	N/A		DRAWINGS:	N/A	
NOTICE OF ANOMOLY:	N/A				
AUTHORIZED SIGNATURE:					
TITLE:	Sr. QS/RC Analyst		DATE:	11/06/15	

Form 1702.3NA  
 Revision: A  
 Effective Date: 08/18/09



Formerly Wyle Laboratories, Inc.

December 14, 2015

**Homewood Products Corporation**  
820 Washington Boulevard  
Pittsburgh, PA 15206

**Attention:** QA Manager

**Subject:** Pursuant to 10 CFR Part 21, this letter notifies Homewood Products Corporation of the existence of a reportable defect and its evaluation.

**Reference:** Homewood Products Corporation Purchase Order No. N61979, NTS Project Number PR029808

The defect is an error in reporting the variability in the dose delivered or lack thereof on the Isomedix Certificate of Processing for the Whippany, NJ facility. The U.S. Nuclear Regulatory Commission (NRC) under 10 CFR Part 50, Appendix B issued a Notice of Nonconformance 99901145/2014-201-01 to Steris stating that the measuring and testing equipment used to determine the applied radiation dose reported on the Isomedix Certificate of Processing provided with each gamma irradiation run did not account for all the uncertainties involved (i.e. density of unrelated products in carriers, off-carrier locations within the irradiator, and Cobalt-60 source decay) and therefore the actual radiation dose applied to components could be less than requested as reported on the Certificate of Processing.

Steris Isomedix Services completed an evaluation of the dose rate variability of items processed in off-carrier locations in the irradiator. Steris Isomedix Services Position Paper dated 4/27/15 states the overall variability (uncertainty) associated with gamma radiation exposures at their Whippany NJ facility.

Steris provided revised Component Irradiation Certification(s) that shows a revised minimum total delivered dose of 32.98 Mrad for Steris Process Run ID 68005A (note the previous cert stated process run number 678005A which was incorrect). The revised Steris certificate is attached.

Per NTS Report PR029808-TR-14, the stated required service condition dose is 30.0 Mrad.

The revised minimum total delivered dose envelopes the stated service condition. Qualification of the component is maintained with reduced margin.

If you have any questions, please feel free to contact me by phone (256) 716-4358, fax (256) 830-2109, or email [ronda.fisher@nts.com](mailto:ronda.fisher@nts.com).

Sincerely,

**National Technical Systems Huntsville**  
Formerly Wyle Laboratories, Inc.

Ronda L. Fisher  
Contracts Administrator II



Formerly Wyle Laboratories, Inc.

December 14, 2015

STERIS

**COMPONENT IRRADIATION CERTIFICATION**

Prepared for: National Technical Systems

Source Type: Cobalt 60 Gamma P.O. # PRPO016016-2

Air Equiv. Required Dose (MRADS)		36.40	
Rate Not to Exceed (MRADS / Hr.)		1.00	
<b>SPECIMENS:</b>			
Qty	Pan No.	Serial No.	Description
1	N/A	N/A	10 HP Electrical Motor
<b>DATA:</b>			
Total Delivered Dose (Air) MRADS:		Min	32.98
		Max	40.86
Dose Rate (Air) MRADS / Hr:		Min	0.647
		Max	0.869
Total Exposure Hours:		42.98	
Static Rotation	Z - WAY:	v	Turntable Rotation
			None:
Date In:	06/25/14	Date Out:	06/28/14
<b>DOSIMETRY:</b>			
Dosimeter Type	Harwell Red 4024 Perspex	Batch	MW
Calibration Date	12/10/13	Calibration Due Date	12/10/14
Readout Instrument:	Beckman DU-640	Serial No.:	4324039
Calibration Date	08/07/13	Calibration Due Date	08/07/14
Comments: Process Run ID: 65005A / "Dose has been adjusted for a process variability of 9.4% in accordance with STERIS Isomedix Services Position Paper Appendix C (Rev 4/27/13)" / Revised Copy 11/13/15			
<b>ATTACHMENTS:</b>			
Worksheets:	N/A	Drawings:	N/A
Notice of Anomaly:	N/A		

Approved By: [Signature] Title: Sr. QS/RC Analyst Date: 11/13/15

Processing Location: STERIS Isomedix Services 9 Apollo Drive, Whippany, NJ 07981 Phone: 973-887-2754 Fax: 973-887-6591  
 The product run described above was processed in accordance with STERIS Isomedix Services Quality System requirements and the Customer approved process (wherever STERIS Isomedix Services facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA). STERIS Isomedix Services operates under a quality system which meets the requirements of the FDA QSR and ISO 13485:2003. STERIS Isomedix Services adheres to requirements provided through ANSI/AAMI/ISO 11137:2016.