




## CONVERSATION RECORD

11/17/2015

NAME OF PERSON(S) CONTACTED OR IN CONTACT WITH YOU Walter Kopecky		DATE OF CONTACT 11/17/2015	TYPE OF CONVERSATION <input type="checkbox"/> E-MAIL <input checked="" type="checkbox"/> TELEPHONE <input type="checkbox"/> INCOMING <input checked="" type="checkbox"/> OUTGOING
E-MAIL ADDRESS kopecky@mvp.net		TELEPHONE NUMBER (314) 275-8224	
ORGANIZATION Nevada Regional Medical Center	DOCKET NUMBER(S) 030-38883		
LICENSE NUMBER(S) Pending	CONTROL NUMBER(S) 589197		
SUBJECT Request for Additional Information			
<b>SUMMARY</b> Based on our review of your application dated October 22, 2015, it appeared that your application had not been completely prepared in accordance with the guidance in NUREG 1556, Vol. 9, Rev. 2, "Consolidated Guidance about Materials Licenses: Program-Specific Guidance about Medical Licenses."  1. Please confirm in writing if you will be using PET radionuclides. If you will be using PET materials, please provide shielding calculations for areas where the materials will be used and stored. In addition, you included Cobalt-57 calibration sources. Please note that these sources should not be listed in your application and will not be included in your license.  2. In your application, you did not include information regarding the calibration of your dose calibrator. Please respond with the following: "Equipment used to measure dosages will be calibrated in accordance with nationally recognized standards or the manufacturer's instructions."  <b>Continue on Page 2</b>			
<b>ACTION REQUIRED (IF ANY)</b> Please submit your response by November 30, 2015 and reference it to my attention as "additional information to control number 589197" to facilitate proper handling in our office. Your response must be currently dated and signed. If you have any questions or require clarification of any of the information stated above, please do not hesitate to contact me at 630-829-9607  In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter and its enclosure will be available electronically for public inspection in the NRC Public Document Room or from the NRC's Agencywide Documents Access and Management System (ADAMS), accessible from the NRC Web site at <a href="http://www.nrc.gov/reading-rm/adams.html">http://www.nrc.gov/reading-rm/adams.html</a> .  <b>Continue on Page 3</b>			
NAME OF PERSON DOCUMENTING CONVERSATION Jennifer L. Bishop			
SIGNATURE 			

**CONVERSATION RECORD (continued)**

SUMMARY: (Continued from page 1)

3. In your application, you did not include information regarding the calibration of your survey instruments. Please response with the following: "Radiation monitoring instruments will be calibrated by a person qualified to perform survey meter calibrations." and/or "We have developed and will implement and maintain written survey meter calibration procedures in accordance with the requirements in 10 CFR 20.1501 and that meet the requirements of 10 CFR 35.61."

Please also include a description of the instrumentation (e.g., gamma counter, solid state detector, portable or stationary count rate meter, portable or stationary dose rate or exposure rate meter, single or multichannel analyzer, liquid scintillation counter, proportional counter) that will be used to perform required surveys.

Please also respond with: "We reserve the right to upgrade our survey instruments as necessary as long as they are adequate to measure the type and level of radiation for which they are used."