



UNITED STATES
NUCLEAR REGULATORY COMMISSION
 REGION IV
 1600 EAST LAMAR BLVD
 ARLINGTON, TEXAS 76011-4511

EMAIL



Name: Dr. Richard Graham License: 05-14892-02
 Radiation Safety Officer Docket: 030-08754
Organization: U.S. Environmental Protection Agency Control: 588333
 Region 8
Phone: 303-312-7080
E-mail Address: graham.richardv@epa.gov
From: Jacqueline D. Cook
Date: January 19, 2016
Subject: Application dated June 29, 2015, for License Renewal
Pages: 2

Dr. Graham:

Per your application dated June 29, 2015, the items on the next page requires your response or clarification. I apologize for the short turnaround time but please respond to this e-mail by close of business Thursday, January 21, 2016. If you need more time to respond, please let me know of an alternate date to respond.

Our fax number is (817) 200-1263. You may respond by e-mail in pdf format if you'd like. My email address is Jackie.Cook@nrc.gov. When responding to this e-mail, please include the license, docket and control numbers located at the top of this page.

Thanking you in advance for your cooperation, assistance, and prompt response in this matter.

/RA/
 Jacqueline D. Cook
 Senior Health Physicist

PUBLIC
 Immediate Release
 Normal Release

NON-PUBLIC
 A.3 Sensitive-Security Related
 A.7 Sensitive Internal
 Other: _____

Reviewer: JAC Date: 1/19/16

1. Please note that "general licensed" check sources are not specifically line listed on your specific license.

Please confirm your understanding of the above.

2. Please note that the NRC does not list an "Alternate Radiation Protection Officer" in a license. The Radiation Safety Officer (RSO) is responsible for the overall management of the radiation safety program; identifying radiation safety problems, initiating, recommending or providing corrective actions; verifying implementation of corrective actions; ensuring compliance with the Commission's rules and regulations and conditions of the license. An RSO should actively participate in the day-to-day management of the radiation safety program. Although the performance of certain radiation safety tasks may be delegated by the RSO to other individuals (such as an alternate Radiation Safety Officer), the responsibility for the overall effectiveness of the radiation safety program and for compliance with NRC rules and regulations and conditions of the license reside with the RSO.

Please confirm your understanding of the above.

3. With the issuance of amendment number 27 to license number 05-14892-02, dated November 12, 2014, please note that we reviewed your current gas chromatograph inventory and based on your statement and a verification of the corresponding sealed source and device registration certificates determined that you are no longer in possession of specifically licensed gas chromatographs. Therefore, the authorization for the use of Ni-63 in specifically licensed gas chromatographs has been removed from the license with the issuance of amendment number 27.

Please confirm this statement, as appropriate.

PUBLIC

- Immediate Release
- Normal Release

NON-PUBLIC

- A.3 Sensitive-Security Related
- A.7 Sensitive Internal
- Other: _____

Reviewer: JAC Date: 4/9/16