



VALLEY INSPECTION
SERVICE INC.

January 8, 2016

U. S. Nuclear Regulatory Commission
Region I, Division of Nuclear Materials Safety
2100 Renaissance Blvd
King of Prussia, PA 19406-2713
(610) 337-5076 voice
Attn: Robin Elliott

Docket No. 03031172
License No. 37-28385-01
Control No. 588973

Ms. Elliott

In response to your request of additional information see below.

1. All calibrations will be performed within the state of PA, under PA license PA-1186
2. With regard to section 6 part A and part B
I am submitting a revised Radiographer Qualification and training procedure TR-100 Rev 2 dated 12/28/2015. All training and testing will be performed by the RSO for Trainee & radiographer assistant. The additional training for radiographer qualification will be sourced out to a 3rd party agency
3. I confirm that the classroom instruction of our training materials includes proper use of personnel monitoring equipment, limitations of survey meters, survey techniques and the levels of radiation from licensed material as required by 10 CFR 34.43(g).
- 4a. Revised sec 6.2 page 15 of 52 to include personal dosimetry -attached
4b. Revised sec 6.1 page 14 of 52 to only RSO or a technician under supervision of the RSO. - attached.
4c. Attached a copy of our calibration stickers for survey meters. All survey meters are adjusted to within +/- 20%. If they cannot be adjusted to within tolerance they are removed from service and sent out for repair.
- 5a. Section 6.3 correction was made to dosimeter page 18 of 52 attached
5b. Section 6.4 correction was made to rate alarm page 20 of 52 attached
5c. Section 8.2 added daily checks for pocket dosimeters and rate alarm meters page 23 of 52 attached

Thank you
Robert W. Kratzer

President & Radiation Safety Officer
Valley Inspection Service, Inc.
430 Business Park Lane
Allentown, PA 18109