

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: Delair Gardi, M.D. 4160 John R Detroit, Michigan 48201 REPORT NUMBER(S) 16-001	2. NRC/REGIONAL OFFICE Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352
--	---

3. DOCKET NUMBER(S) 030-38266	4. LICENSE NUMBER(S) 21-32779-01	5. DATE(S) OF INSPECTION January 14, 2016
--------------------------------------	---	--

LICENSEE:
 The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

Non-cited violation(s) were discussed involving the following requirement(s):

- 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
(Violations and Corrective Actions)

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE	DATE
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Navid Tehrani	<i>Navid Tehrani</i>	1/14/16
BRANCH CHIEF	Aaron McCraw	<i>Aaron McCraw for AIM</i>	1/25/16

Docket File Information

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: Delair Gardi, M.D. 4160 John R Suite 724 Detroit, Michigan 48201 REPORT NUMBER(S) 16-001	2. NRC/REGIONAL OFFICE Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352
---	---

3. DOCKET NUMBER(S) 030-38266	4. LICENSE NUMBER(S) 21-32779-01	5. DATE(S) OF INSPECTION January 14, 2016
--------------------------------------	---	--

6. INSPECTION PROCEDURES USED 87130	7. INSPECTION FOCUS AREAS 03.01-03.07
--	--

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S) 02201	2. PRIORITY 5	3. LICENSEE CONTACT Cari Dzanbazoff (MPC), RSO	4. TELEPHONE NUMBER (734) 662-3197
---------------------------------	----------------------	---	---

Main Office Inspection Next Inspection Date: _____

Field Office Inspection _____

Temporary Job Site Inspection _____

PROGRAM SCOPE

This was an unannounced routine inspection of a private clinic that performed approximately 30 diagnostic nuclear medicine procedures per week. One full time nuclear medicine technologist performed all patient procedures, supplemented by three part time technologists. As of the date of the inspection, the licensee did not conduct nuclear studies on Tuesdays. The licensee obtained licensed material as unit doses from two nuclear pharmacies, and did not use xenon-133, bulk doses, or molybdenum/technetium generators. The licensee performed exclusively cardiac stress and MUGA tests.

The inspector observed one resting and one stress dose of technicium-99m being administered during the inspection. These observations, combined with interviews of available staff, revealed an adequate level of understanding of emergency and material handling procedures and techniques. Dose calibrator constancy checks, package receipt, daily surveys, and waste handling and disposal procedures were successfully demonstrated. An outside consultant performed quarterly program audits that were adequate to oversee the program, and also served as the Radiation Safety Officer (RSO).

Licensed material was adequately secured and not readily accessible to members of the general public. The licensee possessed two radiation survey meters that were calibrated, operational, and performed well in side-by-side comparison with an NRC instrument.

Independent measurements did not indicate readings in excess of Title 10 of the Code of Federal Regulations (10 CFR) Part 20 limits in restricted or unrestricted areas. Personal whole body and extremity dosimetry were observed worn by the staff during the inspection. Dosimetry records did not indicate doses in excess of 10 CFR Part 20 limits for calendar year 2015.

No violations were identified during this inspection.