

## Gauge Inventory as of 1/1/2016

<u>Model</u>	<u>Serial #</u>	<u>Storage Location</u>
4640	T1426	300 Armstrong Rd, Pittston PA 18640
4640	T2691	300 Armstrong Rd, Pittston PA 18640
4640	T2383	300 Armstrong Rd, Pittston PA 18640
4640	T1843	1001 Paxton St, Harrisburg PA 17104
4640	T1851	1061 Manheim Pike, Lancaster PA 17601
4640	T2425	1061 Manheim Pike, Lancaster PA 17601
3440	27263	1001 Paxton St, Harrisburg PA 17104
3440	24943	1001 Paxton St, Harrisburg PA 17104
3450	Z841	300 Armstrong Rd, Pittston PA 18640
3450	Z919	1061 Manheim Pike, Lancaster PA 17601



**North East Technical Services, Inc.**  
75 Aileron Ct., Suite 4  
Westminster, MD 21157  
Ph: 410.751.5090 Fax: 410.751.5091

*Pennsy Supply, Inc. - North*  
30 Small Mountain Road  
Wapwallopen, PA 18660  
ATTN: Steven Grimm

*Shipping Address: 300 Armstrong Road*  
Pittston, PA 18640

## **LEAK TEST CERTIFICATE**

**MD Materials License # MD-13-020-01**

*This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.*

<b>Gauge Model</b>	<b>4640</b>	<b>Gauge S/N</b>	<b>T2383</b>	<b>Leak Test Date</b>	<b>11/25/2015</b>
<b>Source</b>		<b>Reading in microCuries</b>			
	750-7274		0.00006340		

*Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.*

**Reviewed by:** \_\_\_\_\_

**Date:** NOV 30 2015



**North East Technical Services, Inc.**

75 Aileron Ct., Suite 4

Westminster, MD 21157

Ph: 410.751.5090 Fax: 410.751.5091

*Pennsy Supply, Inc. - North*

30 Small Mountain Road

Wapwallopen, PA 18660

ATTN: Steven Grimm

**Shipping Address:** 300 Armstrong Road

Pittston, PA 18640

## **LEAK TEST CERTIFICATE**

**MD Materials License # MD-13-020-01**

*This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.*

<b>Gauge Model</b>	<b>3450</b>	<b>Gauge S/N</b>	<b>Z841</b>	<b>Leak Test Date</b>	<b>11/25/2015</b>
<b>Source</b>			<b>Reading in microCuries</b>		
	751-789		0.00007109		
	47-30164		0.00000		

*Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.*

Reviewed by: \_\_\_\_\_

Date: **NOV 30 2015**



**North East Technical Services, Inc.**  
75 Aileron Ct., Suite 4  
Westminster, MD 21157  
Ph: 410.751.5090 Fax: 410.751.5091

*Pennsy Supply, Inc. - North*  
30 Small Mountain Road  
Wapwallopen, PA 18660  
ATTN: Steven Grimm

*Shipping Address: 300 Armstrong Road*  
Pittston, PA 18640

## ***LEAK TEST CERTIFICATE***

**MD Materials License # MD-13-020-01**

***This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.***

<b>Gauge Model</b>	<b>4640</b>	<b>Gauge S/N</b>	<b>T2691</b>	<b>Leak Test Date</b>	<b>11/25/2015</b>
<b>Source</b>		<b>Reading in microCuries</b>			
	77-5253		0.00006340		

***Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.***

**Reviewed by:** \_\_\_\_\_

**Date:** NOV 30 2015



**North East Technical Services, Inc.**

75 Aileron Ct., Suite 4

Westminster, MD 21157

Ph: 410.751.5090 Fax: 410.751.5091

*Pennsy Supply, Inc. - North*

30 Small Mountain Road

Wapwallopen, PA 18660

ATTN: Steven Grimm

*Shipping Address: 300 Armstrong Road*

*Pittston, PA 18640*

## **LEAK TEST CERTIFICATE**

**MD Materials License # MD-13-020-01**

*This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.*

<b>Gauge Model</b>	<b>4640</b>	<b>Gauge S/N</b>	<b>T1426</b>	<b>Leak Test Date</b>	<b>11/25/2015</b>
<b>Source</b>		<b>Reading in microCuries</b>			
75-5414		0.00006340			

*Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.*

Reviewed by: 

Date: NOV 30 2015



**North East Technical Services, Inc.**  
75 Aileron Ct., Suite 4  
Westminster, MD 21157  
Ph: 410.751.5090 Fax: 410.751.5091

*Pennsy Supply, Inc.*  
1 Clear Spring Road  
Annville, PA 17003  
ATTN: Chris Barton

**Shipping Address:** 1 Clear Spring Road  
Annville, PA 17003

## ***LEAK TEST CERTIFICATE***

**MD Materials License # MD-13-020-01**

***This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.***

<b>Gauge Model</b>	<b>4640</b>	<b>Gauge S/N</b>	<b>T1851</b>	<b>Leak Test Date</b>	<b>7/10/2015</b>
<b>Source</b>		<b>Reading in microCuries</b>			
	75-9924		0.00006145		

***Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.***

**Reviewed by:** \_\_\_\_\_

*Doyle Sinner*

**Date:** \_\_\_\_\_

**JUL 16 2015**

---

**Tuesday, July 21, 2015**





*North East Technical Services, Inc.*  
75 Aileron Ct., Suite 4  
Westminster, MD 21157  
Ph: 410.751.5090 Fax: 410.751.5091

*Pennsy Supply, Inc.*  
1 Clear Spring Road  
Annville, PA 17003  
ATTN: Chris Barton

*Shipping Address: 1 Clear Spring Road  
Annville, PA 17003*

## **LEAK TEST CERTIFICATE**

**MD Materials License # MD-13-020-01**

*This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.*

<b>Gauge Model</b>	<b>4640</b>	<b>Gauge S/N</b>	<b>T2425</b>	<b>Leak Test Date</b>	<b>7/10/2015</b>
<b>Source</b>		<b>Reading in microCuries</b>			
	750-7789		0.00006145		

*Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.*

Reviewed by:                     *Doug La Sierra*                    

Date:                     JUL 16 2015





*North East Technical Services, Inc.*

*75 Aileron Ct., Suite 4*

*Westminster, MD 21157*

*Ph: 410.751.5090 Fax: 410.751.5091*

*Pennsy Supply, Inc.*

*1 Clear Spring Road*

*Annville, PA 17003*

*ATTN: Chris Barton*

*Shipping Address: 1 Clear Spring Road*

*Annville, PA 17003*

## ***LEAK TEST CERTIFICATE***

*MD Materials License # MD-13-020-01*


*This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.*

<b>Gauge Model</b>	<b>4640</b>	<b>Gauge S/N</b>	<b>T1843</b>	<b>Leak Test Date</b>	<b>7/6/2015</b>
	<b>Source</b>		<b>Reading in microCuries</b>		
	75-9916		0.00007489		

*Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.*

*Reviewed by:*           *Douglas S. Simon*          

*Date:*           JUL 16 2015



**North East Technical Services, Inc.**  
 75 Aileron Ct., Suite 4  
 Westminster, MD 21157  
 Ph: 410.751.5090 Fax: 410.751.5091

*Pennsy Supply, Inc.*  
 1 Clear Spring Road  
 Annville, PA 17003  
 ATTN: Chris Barton

**Shipping Address:** 1 Clear Spring Road  
 Annville, PA 17003

## **LEAK TEST CERTIFICATE**

**MD Materials License # MD-13-020-01**

*This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.*

<b>Gauge Model</b>	<b>3440</b>	<b>Gauge S/N</b>	<b>27263</b>	<b>Leak Test Date</b>	<b>7/6/2015</b>
	<b>Source</b>		<b>Reading in microCuries</b>		
	750-1025		0.00007489		
	47-23889		0.00000		

*Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.*

**Reviewed by:**           *Doyle Simon*          

**Date:**           JUL 16 2015



**North East Technical Services, Inc.**  
 75 Aileron Ct., Suite 4  
 Westminster, MD 21157  
 Ph: 410.751.5090 Fax: 410.751.5091

**Pennsy Supply, Inc.**  
 1 Clear Spring Road  
 Annville, PA 17003  
 ATTN: Chris Barton

**Shipping Address:** 1 Clear Spring Road  
 Annville, PA 17003

## **LEAK TEST CERTIFICATE**

**MD Materials License # MD-13-020-01**

*This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.*

<b>Gauge Model</b>	<b>3440</b>	<b>Gauge S/N</b>	<b>24943</b>	<b>Leak Test Date</b>	<b>7/6/2015</b>
	<b>Source</b>		<b>Reading in microCuries</b>		
	75-7084		0.00007489		
	47-21147		0.00000		

*Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.*

**Reviewed by:**           *Donna Sines*          

**Date:**           JUL 16 2015